

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031824\
 Data File : P0102565.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2024 19:56
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO031824

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 06:47:47 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 06:45:40 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.267	3.382	212.8E6	143.0E6	50.994	50.961
2)	SA Decachlor...	9.809	8.249	126.6E6	87071693	50.390	51.001
Target Compounds							
21)	L5 AR-1248-1	5.411	4.434	35758199	23683082	507.991	504.239
22)	L5 AR-1248-2	5.681	4.664	59372050	38288869	490.430	506.060
23)	L5 AR-1248-3	5.881	4.704	62315855	39724042	513.455	499.161
24)	L5 AR-1248-4	6.280	4.872	57417305	45251602	494.440	503.110
25)	L5 AR-1248-5	6.318	5.256	55089892	35327216	492.591	497.413



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031824\
Data File : PO102565.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2024 19:56
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO031824

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 06:47:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031824.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 06:45:40 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.267 min
Delta R.T.: 0.000 min
Response: 212818720
Conc: 50.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#1 Tetrachloro-m-xylene

R.T.: 3.382 min
Delta R.T.: 0.000 min
Response: 142954675
Conc: 50.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.809 min
Delta R.T.: 0.000 min
Response: 126571139
Conc: 50.39 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.249 min
Delta R.T.: 0.000 min
Response: 87071693
Conc: 51.00 ng/ml



#21 AR-1248-1


R.T.: 5.411 min
Delta R.T.: 0.000 min
Response: 35758199
Conc: 507.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#21 AR-1248-1

R.T.: 4.434 min
Delta R.T.: 0.000 min
Response: 23683082
Conc: 504.24 ng/ml




#22 AR-1248-2

R.T.: 5.681 min
Delta R.T.: 0.000 min
Response: 59372050
Conc: 490.43 ng/ml



#22 AR-1248-2


R.T.: 4.664 min
Delta R.T.: 0.000 min
Response: 38288869
Conc: 506.06 ng/ml



#23 AR-1248-3


R.T.: 5.881 min
Delta R.T.: -0.001 min
Response: 62315855
Conc: 513.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#23 AR-1248-3

R.T.: 4.704 min
Delta R.T.: 0.000 min
Response: 39724042
Conc: 499.16 ng/ml




#24 AR-1248-4

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 57417305
Conc: 494.44 ng/ml



#24 AR-1248-4


R.T.: 4.872 min
Delta R.T.: 0.000 min
Response: 45251602
Conc: 503.11 ng/ml



#25 AR-1248-5

R.T.: 6.318 min
Delta R.T.: 0.000 min
Response: 55089892
Conc: 492.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824



#25 AR-1248-5

R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 35327216
Conc: 497.41 ng/ml