

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031824\
 Data File : P0102566.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2024 20:13
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO031824

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 07:14:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 07:12:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.267	3.382	208.6E6	140.0E6	50.432	50.430
2)	SA Decachlor...	9.809	8.247	124.8E6	85967275	50.086	50.471
Target Compounds							
26)	L6 AR-1254-1	6.253	5.217	67262473	60748207	500.963	504.204
27)	L6 AR-1254-2	6.469	5.361	96960115	55335729	502.309	502.279
28)	L6 AR-1254-3	6.831	5.757	91670824	80216702	497.343	504.682
29)	L6 AR-1254-4	7.114	5.983	53640291	39611567	490.391	503.620
30)	L6 AR-1254-5	7.527	6.394	67355119	70036199	494.181	498.747



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031824\
Data File : PO102566.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2024 20:13
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO031824

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 07:14:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031824.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 07:12:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.267 min
Delta R.T.: -0.001 min
Response: 208589759
Conc: 50.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#1 Tetrachloro-m-xylene

R.T.: 3.382 min
Delta R.T.: 0.000 min
Response: 140004989
Conc: 50.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.809 min
Delta R.T.: 0.000 min
Response: 124767186
Conc: 50.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.247 min
Delta R.T.: -0.003 min
Response: 85967275
Conc: 50.47 ng/ml



#26 AR-1254-1


R.T.: 6.253 min
Delta R.T.: -0.001 min
Response: 67262473
Conc: 500.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#26 AR-1254-1

R.T.: 5.217 min
Delta R.T.: 0.000 min
Response: 60748207
Conc: 504.20 ng/ml




#27 AR-1254-2

R.T.: 6.469 min
Delta R.T.: 0.000 min
Response: 96960115
Conc: 502.31 ng/ml



#27 AR-1254-2


R.T.: 5.361 min
Delta R.T.: 0.000 min
Response: 55335729
Conc: 502.28 ng/ml



#28 AR-1254-3


R.T.: 6.831 min
Delta R.T.: -0.001 min
Response: 91670824
Conc: 497.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824




#28 AR-1254-3

R.T.: 5.757 min
Delta R.T.: -0.001 min
Response: 80216702
Conc: 504.68 ng/ml




#29 AR-1254-4

R.T.: 7.114 min
Delta R.T.: 0.000 min
Response: 53640291
Conc: 490.39 ng/ml



#29 AR-1254-4


R.T.: 5.983 min
Delta R.T.: -0.001 min
Response: 39611567
Conc: 503.62 ng/ml



#30 AR-1254-5

R.T.: 7.527 min
Delta R.T.: -0.002 min
Response: 67355119
Conc: 494.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031824



#30 AR-1254-5

R.T.: 6.394 min
Delta R.T.: 0.000 min
Response: 70036199
Conc: 498.75 ng/ml