

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031920\
 Data File : P0067323.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Mar 2020 10:44
 Operator : DD\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 12:30:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 19 12:29:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.941	3.952	1053987	1717053	50.000	50.000
2)	SA Decachlor...	10.924	9.250	1404313	1875801	50.000	50.000
Target Compounds							
3)	L1 AR-1016-1	6.260	5.210	445133	735499	500.000	500.000
4)	L1 AR-1016-2	6.285	5.231	611882	972349	500.000	500.000
5)	L1 AR-1016-3	6.350	5.422	382638	543467	500.000	500.000
6)	L1 AR-1016-4	6.457	5.474	311826	448709	500.000	500.000
7)	L1 AR-1016-5	6.772	5.702	317937	577298	500.000	500.000
31)	L7 AR-1260-1	7.944	6.800	598444	1032071	500.000	500.000
32)	L7 AR-1260-2	8.208	6.997	734988	1271229	500.000	500.000
33)	L7 AR-1260-3	8.571	7.152	573278	1176913	500.000	500.000
34)	L7 AR-1260-4	8.801	7.635	674038	1027130	500.000	500.000
35)	L7 AR-1260-5	9.124	7.882	1403528	2422821	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031920\
Data File : PO067323.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Mar 2020 10:44
Operator : DD\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 12:30:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031920.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 19 12:29:31 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.941 min
Delta R.T.: 0.000 min
Response: 1053987
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.952 min
Delta R.T.: 0.000 min
Response: 1717053
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.924 min
Delta R.T.: 0.000 min
Response: 1404313
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.250 min
Delta R.T.: 0.000 min
Response: 1875801
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 445133
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 5.210 min
Delta R.T.: 0.000 min
Response: 735499
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 6.285 min
Delta R.T.: 0.000 min
Response: 611882
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 5.231 min
Delta R.T.: 0.000 min
Response: 972349
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 6.350 min
Delta R.T.: 0.000 min
Response: 382638
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 5.422 min
Delta R.T.: 0.000 min
Response: 543467
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 6.457 min
Delta R.T.: 0.000 min
Response: 311826
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 5.474 min
Delta R.T.: 0.000 min
Response: 448709
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.772 min
Delta R.T.: 0.000 min
Response: 317937
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 577298
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.944 min
Delta R.T.: 0.000 min
Response: 598444
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.800 min
Delta R.T.: 0.000 min
Response: 1032071
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 8.208 min
Delta R.T.: 0.000 min
Response: 734988
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 6.997 min
Delta R.T.: 0.000 min
Response: 1271229
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 8.571 min
Delta R.T.: 0.000 min
Response: 573278
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 7.152 min
Delta R.T.: 0.000 min
Response: 1176913
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.801 min
Delta R.T.: 0.000 min
Response: 674038
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#34 AR-1260-4

R.T.: 7.635 min
Delta R.T.: 0.000 min
Response: 1027130
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 9.124 min
Delta R.T.: 0.000 min
Response: 1403528
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.882 min
Delta R.T.: 0.000 min
Response: 2422821
Conc: 500.00 ng/ml