

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031920\
 Data File : P0067338.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Mar 2020 16:45
 Operator : DD\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO031920

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 17:00:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 19 16:36:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.956	3.957	1068688	1692829	50.282	48.424
2) SA Decachlor...	10.938	9.253	1432391	1920786	49.852	49.243
Target Compounds						
16) L4 AR-1242-1	6.275	5.215	335864	547979	491.681	469.466
17) L4 AR-1242-2	6.299	5.236	462148	718296	497.482	469.072
18) L4 AR-1242-3	6.364	5.426	289336	403911	488.328	478.738
19) L4 AR-1242-4	6.471	5.524	232412	396008	486.426	468.352
20) L4 AR-1242-5	7.251	6.087	275700	505003	506.319	472.237



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031920\
Data File : P0067338.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Mar 2020 16:45
Operator : DD\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO031920

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 17:00:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031920.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 19 16:36:53 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.956 min
Delta R.T.: 0.016 min
Response: 1068688
Conc: 50.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031920




#1 Tetrachloro-m-xylene

R.T.: 3.957 min
Delta R.T.: 0.005 min
Response: 1692829
Conc: 48.42 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.938 min
Delta R.T.: 0.015 min
Response: 1432391
Conc: 49.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.253 min
Delta R.T.: 0.005 min
Response: 1920786
Conc: 49.24 ng/ml



#16 AR-1242-1


R.T.: 6.275 min
Delta R.T.: 0.015 min
Response: 335864
Conc: 491.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031920




#16 AR-1242-1

R.T.: 5.215 min
Delta R.T.: 0.006 min
Response: 547979
Conc: 469.47 ng/ml




#17 AR-1242-2

R.T.: 6.299 min
Delta R.T.: 0.009 min
Response: 462148
Conc: 497.48 ng/ml



#17 AR-1242-2


R.T.: 5.236 min
Delta R.T.: 0.006 min
Response: 718296
Conc: 469.07 ng/ml



#18 AR-1242-3


R.T.: 6.364 min
Delta R.T.: 0.014 min
Response: 289336
Conc: 488.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031920




#18 AR-1242-3

R.T.: 5.426 min
Delta R.T.: 0.006 min
Response: 403911
Conc: 478.74 ng/ml




#19 AR-1242-4

R.T.: 6.471 min
Delta R.T.: 0.015 min
Response: 232412
Conc: 486.43 ng/ml



#19 AR-1242-4


R.T.: 5.524 min
Delta R.T.: 0.006 min
Response: 396008
Conc: 468.35 ng/ml



#20 AR-1242-5

R.T.: 7.251 min
Delta R.T.: 0.014 min
Response: 275700
Conc: 506.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO031920



#20 AR-1242-5

R.T.: 6.087 min
Delta R.T.: 0.006 min
Response: 505003
Conc: 472.24 ng/ml