

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032024\
 Data File : P0102612.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Mar 2024 09:29
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 20 22:46:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 08:56:28 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.265	3.380	225.0E6	147.1E6	51.035	51.112
2) SA Decachlor...	9.805	8.246	136.7E6	93111340	49.242	51.324
Target Compounds						
16) L4 AR-1242-1	5.409	4.430	52242969	32751483	579.871	561.138
17) L4 AR-1242-2	5.431	4.448	75143923	46921677	569.305	575.451
18) L4 AR-1242-3	5.492	4.620	47989112	25649232	565.976	569.366
19) L4 AR-1242-4	5.589	4.700	38508613	28015511	574.012	540.522
20) L4 AR-1242-5	6.316	5.213	36072357	32302861	576.133	576.910


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO032024\
Data File : PO102612.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Mar 2024 09:29
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 22:46:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031824.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 08:56:28 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.265 min
Delta R.T.: -0.002 min
Response: 225009787
Conc: 51.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.380 min
Delta R.T.: -0.002 min
Response: 147067275
Conc: 51.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.805 min
Delta R.T.: -0.004 min
Response: 136716436
Conc: 49.24 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.246 min
Delta R.T.: -0.003 min
Response: 93111340
Conc: 51.32 ng/ml



#16 AR-1242-1


R.T.: 5.409 min
Delta R.T.: -0.003 min
Response: 52242969
Conc: 579.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.430 min
Delta R.T.: -0.003 min
Response: 32751483
Conc: 561.14 ng/ml




#17 AR-1242-2

R.T.: 5.431 min
Delta R.T.: -0.003 min
Response: 75143923
Conc: 569.31 ng/ml



#17 AR-1242-2


R.T.: 4.448 min
Delta R.T.: -0.003 min
Response: 46921677
Conc: 575.45 ng/ml



#18 AR-1242-3


R.T.: 5.492 min
Delta R.T.: -0.003 min
Response: 47989112
Conc: 565.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.620 min
Delta R.T.: -0.003 min
Response: 25649232
Conc: 569.37 ng/ml




#19 AR-1242-4

R.T.: 5.589 min
Delta R.T.: -0.002 min
Response: 38508613
Conc: 574.01 ng/ml



#19 AR-1242-4


R.T.: 4.700 min
Delta R.T.: -0.004 min
Response: 28015511
Conc: 540.52 ng/ml



#20 AR-1242-5

R.T.: 6.316 min
Delta R.T.: -0.003 min
Response: 36072357
Conc: 576.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.213 min
Delta R.T.: -0.003 min
Response: 32302861
Conc: 576.91 ng/ml