

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0032122\
 Data File : P0085540.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Mar 2022 14:32
 Operator : YP\AJ
 Sample : N2009-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 21 15:48:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0031522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 16 04:43:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.491	3.590	2365976	546476	11.571	12.755
2)	SA Decachlor...	10.392	8.605	1852053	570335	15.407	14.855
Target Compounds							
3)	L1 AR-1016-1	5.676	4.678	75856	30985	12.054	19.416 #
4)	L1 AR-1016-2	5.701	4.695	114937	51710	12.548	22.349 #
5)	L1 AR-1016-3	5.764	4.870	136770	43927	25.127	35.148 #
6)	L1 AR-1016-4	5.865	4.917	99783	72838	22.621	66.724 #
7)	L1 AR-1016-5	6.164	5.128	267311	80417	62.953	59.454
8)	L2 AR-1221-1	4.701	3.831	809801	-2655	367.122	N.D. #
9)	L2 AR-1221-2	4.774	3.871	3132	133804	1.923	318.656 #
10)	L2 AR-1221-3	4.871	3.968	91992	19364	17.890	15.100
11)	L3 AR-1232-1	4.871	3.968	91992	19364	19.547	16.344
12)	L3 AR-1232-2	5.412	4.695	107739	51710	46.754	47.459
13)	L3 AR-1232-3	5.701	4.870	114937	43927	27.238	72.286 #
14)	L3 AR-1232-4	5.865	4.956	99783	61589	49.693	110.512 #
15)	L3 AR-1232-5	5.954	5.128	346055	80417	235.123	130.425 #
16)	L4 AR-1242-1	5.676	4.678	75856	30985	14.921	23.714 #
17)	L4 AR-1242-2	5.701	4.695	114937	51710	15.484	27.981 #
18)	L4 AR-1242-3	5.764	4.870	136770	43927	31.123	43.443 #
19)	L4 AR-1242-4	5.865	4.956	99783	61589	27.498	59.251 #
20)	L4 AR-1242-5	6.586	5.479	5236055	170663	1473.540	138.286 #
21)	L5 AR-1248-1	5.676	4.678	75856	30985	20.349	32.049 #
22)	L5 AR-1248-2	5.954	4.917	346055	72838	65.287	49.742
23)	L5 AR-1248-3	6.164	4.956	267311	61589	45.435	40.182
24)	L5 AR-1248-4	6.586	5.128	5236055	80417	812.179	45.751 #
25)	L5 AR-1248-5	6.586	5.518	5236055	204237	845.711	118.909 #
26)	L6 AR-1254-1	6.546	5.479	469898	170663	70.904	64.935
27)	L6 AR-1254-2	6.769	5.628	700849	146204	70.031	62.658
28)	L6 AR-1254-3	7.138	6.036	477255	387451	46.374	105.491 #
29)	L6 AR-1254-4	7.426	6.261	445761	101212	63.536	45.692 #
30)	L6 AR-1254-5	7.850	6.680	1294682	394285	169.205	121.215 #
31)	L7 AR-1260-1	7.305	6.163	701125	242793	101.612	102.409
32)	L7 AR-1260-2	7.565	6.352	965103	236998	123.373	84.897 #
33)	L7 AR-1260-3	7.928	6.504	420510	280809	71.593	106.804 #
34)	L7 AR-1260-4	8.162	6.979	1085998	181929	157.231	81.225 #
35)	L7 AR-1260-5	8.488	7.223	1015987	396365	72.317	78.678
36)	L8 AR-1262-1	7.928	6.680	420510	394285	46.559	229.638 #
37)	L8 AR-1262-2	8.488	6.979	1015987	181929	63.764	63.528
38)	L8 AR-1262-3	8.827	7.507	634552	69254	61.392	31.623 #
39)	L8 AR-1262-4	8.887	7.570	384099	307515	81.778	77.059
40)	L8 AR-1262-5	9.595	8.068	242050	-27220	45.515	N.D. #
41)	L9 AR-1268-1	8.827	7.507	634552	69254	32.879	10.428 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032122\
Data File : P0085540.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Mar 2022 14:32
Operator : YP\AJ
Sample : N2009-05
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 15:48:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 04:43:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.887	7.570	384099	307515	22.106	52.175 #
43)	L9 AR-1268-3	9.132	7.784	137131	10034	9.291	2.020 #
44)	L9 AR-1268-4	9.595	8.068	242050	-27220	39.671	N.D. #
45)	L9 AR-1268-5	10.030	8.354	227591	40578	4.714	2.897 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032122\
Data File : P0085540.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Mar 2022 14:32
Operator : YP\AJ
Sample : N2009-05
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 15:48:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 04:43:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.491 min
Delta R.T.: -0.001 min
Response: 2365976
Conc: 11.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.590 min
Delta R.T.: 0.000 min
Response: 546476
Conc: 12.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.392 min
Delta R.T.: -0.004 min
Response: 1852053
Conc: 15.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.605 min
Delta R.T.: -0.002 min
Response: 570335
Conc: 14.86 ng/ml



#3 AR-1016-1


R.T.: 5.676 min
Delta R.T.: -0.004 min
Response: 75856
Conc: 12.05 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.678 min
Delta R.T.: 0.000 min
Response: 30985
Conc: 19.42 ng/ml




#4 AR-1016-2

R.T.: 5.701 min
Delta R.T.: -0.001 min
Response: 114937
Conc: 12.55 ng/ml



#4 AR-1016-2


R.T.: 4.695 min
Delta R.T.: -0.002 min
Response: 51710
Conc: 22.35 ng/ml



#5 AR-1016-3


R.T.: 5.764 min
Delta R.T.: -0.003 min
Response: 136770
Conc: 25.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.870 min
Delta R.T.: -0.002 min
Response: 43927
Conc: 35.15 ng/ml




#6 AR-1016-4

R.T.: 5.865 min
Delta R.T.: -0.001 min
Response: 99783
Conc: 22.62 ng/ml



#6 AR-1016-4


R.T.: 4.917 min
Delta R.T.: 0.001 min
Response: 72838
Conc: 66.72 ng/ml



#7 AR-1016-5


R.T.: 6.164 min
Delta R.T.: -0.001 min
Response: 267311
Conc: 62.95 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.128 min
Delta R.T.: -0.001 min
Response: 80417
Conc: 59.45 ng/ml




#8 AR-1221-1

R.T.: 4.701 min
Delta R.T.: -0.002 min
Response: 809801
Conc: 367.12 ng/ml



#8 AR-1221-1


R.T.: 3.831 min
Delta R.T.: 0.026 min
Response: -2655
Conc: N.D.



#9 AR-1221-2


R.T.: 4.774 min
Delta R.T.: -0.017 min
Response: 3132
Conc: 1.92 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.871 min
Delta R.T.: -0.019 min
Response: 133804
Conc: 318.66 ng/ml




#10 AR-1221-3

R.T.: 4.871 min
Delta R.T.: 0.002 min
Response: 91992
Conc: 17.89 ng/ml



#10 AR-1221-3


R.T.: 3.968 min
Delta R.T.: 0.002 min
Response: 19364
Conc: 15.10 ng/ml



#11 AR-1232-1


R.T.: 4.871 min
Delta R.T.: 0.002 min
Response: 91992
Conc: 19.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.968 min
Delta R.T.: 0.002 min
Response: 19364
Conc: 16.34 ng/ml




#12 AR-1232-2

R.T.: 5.412 min
Delta R.T.: 0.004 min
Response: 107739
Conc: 46.75 ng/ml



#12 AR-1232-2


R.T.: 4.695 min
Delta R.T.: -0.001 min
Response: 51710
Conc: 47.46 ng/ml



#13 AR-1232-3


R.T.: 5.701 min
Delta R.T.: -0.002 min
Response: 114937
Conc: 27.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.870 min
Delta R.T.: -0.002 min
Response: 43927
Conc: 72.29 ng/ml




#14 AR-1232-4

R.T.: 5.865 min
Delta R.T.: -0.002 min
Response: 99783
Conc: 49.69 ng/ml



#14 AR-1232-4


R.T.: 4.956 min
Delta R.T.: 0.000 min
Response: 61589
Conc: 110.51 ng/ml



#15 AR-1232-5


R.T.: 5.954 min
Delta R.T.: -0.004 min
Response: 346055
Conc: 235.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.128 min
Delta R.T.: 0.000 min
Response: 80417
Conc: 130.42 ng/ml




#16 AR-1242-1

R.T.: 5.676 min
Delta R.T.: -0.004 min
Response: 75856
Conc: 14.92 ng/ml



#16 AR-1242-1


R.T.: 4.678 min
Delta R.T.: 0.000 min
Response: 30985
Conc: 23.71 ng/ml



#17 AR-1242-2


R.T.: 5.701 min
Delta R.T.: -0.001 min
Response: 114937
Conc: 15.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.695 min
Delta R.T.: -0.001 min
Response: 51710
Conc: 27.98 ng/ml




#18 AR-1242-3

R.T.: 5.764 min
Delta R.T.: -0.002 min
Response: 136770
Conc: 31.12 ng/ml



#18 AR-1242-3


R.T.: 4.870 min
Delta R.T.: -0.002 min
Response: 43927
Conc: 43.44 ng/ml



#19 AR-1242-4


R.T.: 5.865 min
Delta R.T.: -0.002 min
Response: 99783
Conc: 27.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.956 min
Delta R.T.: 0.000 min
Response: 61589
Conc: 59.25 ng/ml




#20 AR-1242-5

R.T.: 6.586 min
Delta R.T.: -0.027 min
Response: 5236055
Conc: 1473.54 ng/ml



#20 AR-1242-5


R.T.: 5.479 min
Delta R.T.: -0.001 min
Response: 170663
Conc: 138.29 ng/ml



#21 AR-1248-1


R.T.: 5.676 min
Delta R.T.: -0.002 min
Response: 75856
Conc: 20.35 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.678 min
Delta R.T.: 0.000 min
Response: 30985
Conc: 32.05 ng/ml




#22 AR-1248-2

R.T.: 5.954 min
Delta R.T.: -0.002 min
Response: 346055
Conc: 65.29 ng/ml



#22 AR-1248-2


R.T.: 4.917 min
Delta R.T.: 0.001 min
Response: 72838
Conc: 49.74 ng/ml



#23 AR-1248-3


R.T.: 6.164 min
Delta R.T.: 0.000 min
Response: 267311
Conc: 45.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.956 min
Delta R.T.: 0.000 min
Response: 61589
Conc: 40.18 ng/ml




#24 AR-1248-4

R.T.: 6.586 min
Delta R.T.: 0.013 min
Response: 5236055
Conc: 812.18 ng/ml



#24 AR-1248-4


R.T.: 5.128 min
Delta R.T.: 0.000 min
Response: 80417
Conc: 45.75 ng/ml



#25 AR-1248-5


R.T.: 6.586 min
Delta R.T.: -0.026 min
Response: 5236055
Conc: 845.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.518 min
Delta R.T.: -0.003 min
Response: 204237
Conc: 118.91 ng/ml




#26 AR-1254-1

R.T.: 6.546 min
Delta R.T.: 0.000 min
Response: 469898
Conc: 70.90 ng/ml



#26 AR-1254-1


R.T.: 5.479 min
Delta R.T.: 0.000 min
Response: 170663
Conc: 64.93 ng/ml



#27 AR-1254-2


R.T.: 6.769 min
Delta R.T.: 0.000 min
Response: 700849
Conc: 70.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.628 min
Delta R.T.: 0.000 min
Response: 146204
Conc: 62.66 ng/ml




#28 AR-1254-3

R.T.: 7.138 min
Delta R.T.: 0.000 min
Response: 477255
Conc: 46.37 ng/ml



#28 AR-1254-3


R.T.: 6.036 min
Delta R.T.: 0.004 min
Response: 387451
Conc: 105.49 ng/ml



#29 AR-1254-4


R.T.: 7.426 min
Delta R.T.: -0.001 min
Response: 445761
Conc: 63.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.261 min
Delta R.T.: 0.000 min
Response: 101212
Conc: 45.69 ng/ml




#30 AR-1254-5

R.T.: 7.850 min
Delta R.T.: -0.002 min
Response: 1294682
Conc: 169.20 ng/ml



#30 AR-1254-5


R.T.: 6.680 min
Delta R.T.: 0.000 min
Response: 394285
Conc: 121.21 ng/ml



#31 AR-1260-1


R.T.: 7.305 min
Delta R.T.: -0.002 min
Response: 701125
Conc: 101.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.163 min
Delta R.T.: -0.002 min
Response: 242793
Conc: 102.41 ng/ml




#32 AR-1260-2

R.T.: 7.565 min
Delta R.T.: -0.001 min
Response: 965103
Conc: 123.37 ng/ml



#32 AR-1260-2


R.T.: 6.352 min
Delta R.T.: -0.003 min
Response: 236998
Conc: 84.90 ng/ml



#33 AR-1260-3


R.T.: 7.928 min
Delta R.T.: -0.001 min
Response: 420510
Conc: 71.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.504 min
Delta R.T.: -0.004 min
Response: 280809
Conc: 106.80 ng/ml




#34 AR-1260-4

R.T.: 8.162 min
Delta R.T.: 0.002 min
Response: 1085998
Conc: 157.23 ng/ml



#34 AR-1260-4


R.T.: 6.979 min
Delta R.T.: -0.002 min
Response: 181929
Conc: 81.23 ng/ml



#35 AR-1260-5


R.T.: 8.488 min
Delta R.T.: -0.003 min
Response: 1015987
Conc: 72.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.223 min
Delta R.T.: -0.002 min
Response: 396365
Conc: 78.68 ng/ml




#36 AR-1262-1

R.T.: 7.928 min
Delta R.T.: -0.001 min
Response: 420510
Conc: 46.56 ng/ml



#36 AR-1262-1


R.T.: 6.680 min
Delta R.T.: 0.000 min
Response: 394285
Conc: 229.64 ng/ml



#37 AR-1262-2


R.T.: 8.488 min
Delta R.T.: -0.002 min
Response: 1015987
Conc: 63.76 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 6.979 min
Delta R.T.: 0.000 min
Response: 181929
Conc: 63.53 ng/ml




#38 AR-1262-3

R.T.: 8.827 min
Delta R.T.: 0.009 min
Response: 634552
Conc: 61.39 ng/ml



#38 AR-1262-3


R.T.: 7.507 min
Delta R.T.: 0.000 min
Response: 69254
Conc: 31.62 ng/ml



#39 AR-1262-4


R.T.: 8.887 min
Delta R.T.: -0.023 min
Response: 384099
Conc: 81.78 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.570 min
Delta R.T.: 0.000 min
Response: 307515
Conc: 77.06 ng/ml




#40 AR-1262-5

R.T.: 9.595 min
Delta R.T.: -0.002 min
Response: 242050
Conc: 45.52 ng/ml



#40 AR-1262-5


R.T.: 8.068 min
Delta R.T.: 0.000 min
Response: -27220
Conc: N.D.



#41 AR-1268-1


R.T.: 8.827 min
Delta R.T.: 0.013 min
Response: 634552
Conc: 32.88 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.507 min
Delta R.T.: 0.000 min
Response: 69254
Conc: 10.43 ng/ml




#42 AR-1268-2

R.T.: 8.887 min
Delta R.T.: -0.024 min
Response: 384099
Conc: 22.11 ng/ml



#42 AR-1268-2


R.T.: 7.570 min
Delta R.T.: -0.002 min
Response: 307515
Conc: 52.17 ng/ml



#43 AR-1268-3


R.T.: 9.132 min
Delta R.T.: -0.019 min
Response: 137131
Conc: 9.29 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.784 min
Delta R.T.: 0.005 min
Response: 10034
Conc: 2.02 ng/ml




#44 AR-1268-4

R.T.: 9.595 min
Delta R.T.: -0.002 min
Response: 242050
Conc: 39.67 ng/ml



#44 AR-1268-4


R.T.: 8.068 min
Delta R.T.: 0.000 min
Response: -27220
Conc: N.D.



#45 AR-1268-5

R.T.: 10.030 min
Delta R.T.: -0.003 min
Response: 227591
Conc: 4.71 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.354 min
Delta R.T.: 0.000 min
Response: 40578
Conc: 2.90 ng/ml