

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032122\
 Data File : P0085540.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 Mar 2022 14:32
 Operator : YP\AJ
 Sample : N2009-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 31405

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 21 15:48:25 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 16 04:43:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.491	3.590	2365976	546476	11.571	12.755
2)	SA Decachlor...	10.392	8.604	1852053	650195	15.407	16.935m
Target Compounds							
26)	L6 AR-1254-1	6.546	5.479	469898	170663	70.904	64.935
27)	L6 AR-1254-2	6.769	5.628	700849	146204	70.031	62.658
28)	L6 AR-1254-3	7.138	6.036	477255	387451	46.374	105.491 #
29)	L6 AR-1254-4	7.426	6.261	445761	101212	63.536	45.692 #
30)	L6 AR-1254-5	7.850	6.680	924131	394285	120.777m	121.215
31)	L7 AR-1260-1	7.305	6.163	701125	242793	101.612	102.409
32)	L7 AR-1260-2	7.565	6.352	965103	236998	123.373	84.897 #
33)	L7 AR-1260-3	7.928	6.504	420510	280809	71.593	106.804 #
34)	L7 AR-1260-4	8.162	6.979	1085998	181929	157.231	81.225 #
35)	L7 AR-1260-5	8.488	7.223	1015987	396365	72.317	78.678



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032122\
Data File : P0085540.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 Mar 2022 14:32
Operator : YP\AJ
Sample : N2009-05
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
31405

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 15:48:25 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 04:43:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.491 min
Delta R.T.: -0.001 min
Response: 2365976
Conc: 11.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
31405




#1 Tetrachloro-m-xylene

R.T.: 3.590 min
Delta R.T.: 0.000 min
Response: 546476
Conc: 12.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.392 min
Delta R.T.: -0.004 min
Response: 1852053
Conc: 15.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.604 min
Delta R.T.: -0.002 min
Response: 650195
Conc: 16.94 ng/ml m



#26 AR-1254-1


R.T.: 6.546 min
Delta R.T.: 0.000 min
Response: 469898
Conc: 70.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
31405




#26 AR-1254-1

R.T.: 5.479 min
Delta R.T.: 0.000 min
Response: 170663
Conc: 64.93 ng/ml




#27 AR-1254-2

R.T.: 6.769 min
Delta R.T.: 0.000 min
Response: 700849
Conc: 70.03 ng/ml



#27 AR-1254-2


R.T.: 5.628 min
Delta R.T.: 0.000 min
Response: 146204
Conc: 62.66 ng/ml



#28 AR-1254-3


R.T.: 7.138 min
Delta R.T.: 0.000 min
Response: 477255
Conc: 46.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
31405




#28 AR-1254-3

R.T.: 6.036 min
Delta R.T.: 0.004 min
Response: 387451
Conc: 105.49 ng/ml




#29 AR-1254-4

R.T.: 7.426 min
Delta R.T.: -0.001 min
Response: 445761
Conc: 63.54 ng/ml



#29 AR-1254-4


R.T.: 6.261 min
Delta R.T.: 0.000 min
Response: 101212
Conc: 45.69 ng/ml



#30 AR-1254-5


R.T.: 7.850 min
Delta R.T.: -0.002 min
Response: 924131
Conc: 120.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
31405




#30 AR-1254-5

R.T.: 6.680 min
Delta R.T.: 0.000 min
Response: 394285
Conc: 121.21 ng/ml




#31 AR-1260-1

R.T.: 7.305 min
Delta R.T.: -0.002 min
Response: 701125
Conc: 101.61 ng/ml



#31 AR-1260-1


R.T.: 6.163 min
Delta R.T.: -0.002 min
Response: 242793
Conc: 102.41 ng/ml



#32 AR-1260-2


R.T.: 7.565 min
Delta R.T.: -0.001 min
Response: 965103
Conc: 123.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
31405




#32 AR-1260-2

R.T.: 6.352 min
Delta R.T.: -0.003 min
Response: 236998
Conc: 84.90 ng/ml




#33 AR-1260-3

R.T.: 7.928 min
Delta R.T.: -0.001 min
Response: 420510
Conc: 71.59 ng/ml



#33 AR-1260-3


R.T.: 6.504 min
Delta R.T.: -0.004 min
Response: 280809
Conc: 106.80 ng/ml



#34 AR-1260-4


R.T.: 8.162 min
Delta R.T.: 0.002 min
Response: 1085998
Conc: 157.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
31405




#34 AR-1260-4

R.T.: 6.979 min
Delta R.T.: -0.002 min
Response: 181929
Conc: 81.23 ng/ml



#35 AR-1260-5

R.T.: 8.488 min
Delta R.T.: -0.003 min
Response: 1015987
Conc: 72.32 ng/ml



#35 AR-1260-5

R.T.: 7.223 min
Delta R.T.: -0.002 min
Response: 396365
Conc: 78.68 ng/ml