

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032221\
 Data File : P0076445.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Mar 2021 22:02
 Operator : DD\AJ
 Sample : PB135270BSD
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB135270BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 00:42:55 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 16 07:18:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.917	3.930	1240486	814223	21.836	19.516
2)	SA Decachlor...	10.911	9.191	888947	898122	23.111	23.390
Target Compounds							
3)	L1 AR-1016-1	6.240	5.178	849364	570151	515.479	487.163
4)	L1 AR-1016-2	6.264	5.198	1362806	1089770	533.342	494.678
5)	L1 AR-1016-3	6.330	5.388	872023	496280	553.760	482.662
6)	L1 AR-1016-4	6.437	5.438	673423	460802	544.427	469.795
7)	L1 AR-1016-5	6.752	5.665	649575	541822	506.059	476.081
8)	L2 AR-1221-1	5.164	4.186	76118	65793	144.158	149.213
9)	L2 AR-1221-2	5.268	4.286	103695	77892	253.168	238.511
10)	L2 AR-1221-3	5.354	4.374	477753	359840	355.914	341.328
11)	L3 AR-1232-1	5.354	4.374	477753	359840	442.850	421.018
12)	L3 AR-1232-2	5.943	5.198	714775	1089770	1213.950	1154.626
13)	L3 AR-1232-3	6.264	5.388	1362806	496280	1308.175	1194.876
14)	L3 AR-1232-4	6.437	5.482	673423	527969	1379.179	1260.977
15)	L3 AR-1232-5	6.534	5.665	555700	541822	1491.985	1230.650
16)	L4 AR-1242-1	6.240	5.178	849364	570151	672.565	642.635
17)	L4 AR-1242-2	6.264	5.198	1362806	1089770	687.617	629.619
18)	L4 AR-1242-3	6.330	5.388	872023	496280	688.172	616.122
19)	L4 AR-1242-4	6.437	5.482	673423	527969	695.468	608.236
20)	L4 AR-1242-5	7.222	6.042	88140	456028	87.730	452.503 #
21)	L5 AR-1248-1	6.240	5.178	849364	570151	900.918	830.174
22)	L5 AR-1248-2	6.534	5.438	555700	460802	368.895	356.146
23)	L5 AR-1248-3	6.752	5.482	649575	527969	374.521	413.056
24)	L5 AR-1248-4	7.182	5.665	96921	541822	55.090	368.471 #
25)	L5 AR-1248-5	7.222	6.089	88140	77300	49.835	49.982
26)	L6 AR-1254-1	7.150	6.042	494759	456028	241.416	199.530
27)	L6 AR-1254-2	7.380	6.202	520756	432600	167.577	218.424 #
28)	L6 AR-1254-3	7.762	6.637	271329	767547	85.285	254.827 #
29)	L6 AR-1254-4	8.057	6.860	163508	77037	78.042	43.655 #
30)	L6 AR-1254-5	8.487	7.285	1638982	1453386	713.632	555.136
31)	L7 AR-1260-1	7.929	6.755	1254094	1093096	539.388	532.103
32)	L7 AR-1260-2	8.195	6.953	1390287	1314135	544.497	541.858
33)	L7 AR-1260-3	8.561	7.106	910750	1210405	459.404	535.187
34)	L7 AR-1260-4	8.792	7.588	1076297	891628	476.417	462.705
35)	L7 AR-1260-5	9.116	7.836	1884289	2057214	460.982	465.324
36)	L8 AR-1262-1	8.561	7.285	910750	1453386	307.826	995.201 #
37)	L8 AR-1262-2	9.116	7.836	1884289	2057214	389.234	415.297
38)	L8 AR-1262-3	9.447	8.120	1199977	444959	363.096	207.114 #
39)	L8 AR-1262-4	9.498f	8.182	689701	1526081	268.298	403.485 #
40)	L8 AR-1262-5	10.175	8.680	434558	491234	264.178	283.771
41)	L9 AR-1268-1	9.447f	8.120	1199977	444959	205.998	75.763 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032221\
 Data File : P0076445.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Mar 2021 22:02
 Operator : DD\AJ
 Sample : PB135270BSD
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB135270BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 00:42:55 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 16 07:18:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.498f	8.182	689701	1526081	130.012	286.520 #
43) L9	AR-1268-3	9.742	8.390	37933	38741	8.182	8.168
44) L9	AR-1268-4	10.175	8.680	434558	491234	249.346	264.339
45) L9	AR-1268-5	10.582	8.955	159124	170002	11.558	12.738



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO032221\
Data File : PO076445.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Mar 2021 22:02
Operator : DD\AJ
Sample : PB135270BSD
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB135270BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 00:42:55 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 16 07:18:57 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.917 min
Delta R.T.: -0.005 min
Response: 1240486
Conc: 21.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#1 Tetrachloro-m-xylene

R.T.: 3.930 min
Delta R.T.: -0.006 min
Response: 814223
Conc: 19.52 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.911 min
Delta R.T.: -0.013 min
Response: 888947
Conc: 23.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.191 min
Delta R.T.: -0.013 min
Response: 898122
Conc: 23.39 ng/ml



#3 AR-1016-1


R.T.: 6.240 min
Delta R.T.: -0.006 min
Response: 849364
Conc: 515.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#3 AR-1016-1

R.T.: 5.178 min
Delta R.T.: -0.007 min
Response: 570151
Conc: 487.16 ng/ml




#4 AR-1016-2

R.T.: 6.264 min
Delta R.T.: -0.005 min
Response: 1362806
Conc: 533.34 ng/ml



#4 AR-1016-2


R.T.: 5.198 min
Delta R.T.: -0.007 min
Response: 1089770
Conc: 494.68 ng/ml



#5 AR-1016-3


R.T.: 6.330 min
Delta R.T.: -0.005 min
Response: 872023
Conc: 553.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#5 AR-1016-3

R.T.: 5.388 min
Delta R.T.: -0.008 min
Response: 496280
Conc: 482.66 ng/ml




#6 AR-1016-4

R.T.: 6.437 min
Delta R.T.: -0.005 min
Response: 673423
Conc: 544.43 ng/ml



#6 AR-1016-4


R.T.: 5.438 min
Delta R.T.: -0.008 min
Response: 460802
Conc: 469.80 ng/ml



#7 AR-1016-5


R.T.: 6.752 min
Delta R.T.: -0.006 min
Response: 649575
Conc: 506.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#7 AR-1016-5

R.T.: 5.665 min
Delta R.T.: -0.009 min
Response: 541822
Conc: 476.08 ng/ml




#8 AR-1221-1

R.T.: 5.164 min
Delta R.T.: -0.012 min
Response: 76118
Conc: 144.16 ng/ml



#8 AR-1221-1


R.T.: 4.186 min
Delta R.T.: -0.009 min
Response: 65793
Conc: 149.21 ng/ml



#9 AR-1221-2


R.T.: 5.268 min
Delta R.T.: -0.007 min
Response: 103695
Conc: 253.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#9 AR-1221-2

R.T.: 4.286 min
Delta R.T.: -0.007 min
Response: 77892
Conc: 238.51 ng/ml




#10 AR-1221-3

R.T.: 5.354 min
Delta R.T.: -0.006 min
Response: 477753
Conc: 355.91 ng/ml



#10 AR-1221-3


R.T.: 4.374 min
Delta R.T.: -0.007 min
Response: 359840
Conc: 341.33 ng/ml



#11 AR-1232-1


R.T.: 5.354 min
Delta R.T.: -0.007 min
Response: 477753
Conc: 442.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#11 AR-1232-1

R.T.: 4.374 min
Delta R.T.: -0.009 min
Response: 359840
Conc: 421.02 ng/ml




#12 AR-1232-2

R.T.: 5.943 min
Delta R.T.: -0.007 min
Response: 714775
Conc: 1213.95 ng/ml



#12 AR-1232-2


R.T.: 5.198 min
Delta R.T.: -0.010 min
Response: 1089770
Conc: 1154.63 ng/ml



#13 AR-1232-3


R.T.: 6.264 min
Delta R.T.: -0.008 min
Response: 1362806
Conc: 1308.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#13 AR-1232-3

R.T.: 5.388 min
Delta R.T.: -0.010 min
Response: 496280
Conc: 1194.88 ng/ml




#14 AR-1232-4

R.T.: 6.437 min
Delta R.T.: -0.009 min
Response: 673423
Conc: 1379.18 ng/ml



#14 AR-1232-4


R.T.: 5.482 min
Delta R.T.: -0.009 min
Response: 527969
Conc: 1260.98 ng/ml



#15 AR-1232-5


R.T.: 6.534 min
Delta R.T.: -0.008 min
Response: 555700
Conc: 1491.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#15 AR-1232-5

R.T.: 5.665 min
Delta R.T.: -0.010 min
Response: 541822
Conc: 1230.65 ng/ml




#16 AR-1242-1

R.T.: 6.240 min
Delta R.T.: -0.006 min
Response: 849364
Conc: 672.57 ng/ml



#16 AR-1242-1


R.T.: 5.178 min
Delta R.T.: -0.008 min
Response: 570151
Conc: 642.63 ng/ml



#17 AR-1242-2


R.T.: 6.264 min
Delta R.T.: -0.006 min
Response: 1362806
Conc: 687.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#17 AR-1242-2

R.T.: 5.198 min
Delta R.T.: -0.008 min
Response: 1089770
Conc: 629.62 ng/ml




#18 AR-1242-3

R.T.: 6.330 min
Delta R.T.: -0.006 min
Response: 872023
Conc: 688.17 ng/ml



#18 AR-1242-3


R.T.: 5.388 min
Delta R.T.: -0.008 min
Response: 496280
Conc: 616.12 ng/ml



#19 AR-1242-4


R.T.: 6.437 min
Delta R.T.: -0.007 min
Response: 673423
Conc: 695.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#19 AR-1242-4

R.T.: 5.482 min
Delta R.T.: -0.008 min
Response: 527969
Conc: 608.24 ng/ml




#20 AR-1242-5

R.T.: 7.222 min
Delta R.T.: -0.006 min
Response: 88140
Conc: 87.73 ng/ml



#20 AR-1242-5


R.T.: 6.042 min
Delta R.T.: -0.011 min
Response: 456028
Conc: 452.50 ng/ml



#21 AR-1248-1


R.T.: 6.240 min
Delta R.T.: -0.006 min
Response: 849364
Conc: 900.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#21 AR-1248-1

R.T.: 5.178 min
Delta R.T.: -0.007 min
Response: 570151
Conc: 830.17 ng/ml




#22 AR-1248-2

R.T.: 6.534 min
Delta R.T.: -0.006 min
Response: 555700
Conc: 368.90 ng/ml



#22 AR-1248-2


R.T.: 5.438 min
Delta R.T.: -0.008 min
Response: 460802
Conc: 356.15 ng/ml



#23 AR-1248-3


R.T.: 6.752 min
Delta R.T.: -0.006 min
Response: 649575
Conc: 374.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#23 AR-1248-3

R.T.: 5.482 min
Delta R.T.: -0.008 min
Response: 527969
Conc: 413.06 ng/ml




#24 AR-1248-4

R.T.: 7.182 min
Delta R.T.: -0.005 min
Response: 96921
Conc: 55.09 ng/ml



#24 AR-1248-4


R.T.: 5.665 min
Delta R.T.: -0.008 min
Response: 541822
Conc: 368.47 ng/ml



#25 AR-1248-5


R.T.: 7.222 min
Delta R.T.: -0.005 min
Response: 88140
Conc: 49.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#25 AR-1248-5

R.T.: 6.089 min
Delta R.T.: -0.007 min
Response: 77300
Conc: 49.98 ng/ml




#26 AR-1254-1

R.T.: 7.150 min
Delta R.T.: -0.006 min
Response: 494759
Conc: 241.42 ng/ml



#26 AR-1254-1


R.T.: 6.042 min
Delta R.T.: -0.009 min
Response: 456028
Conc: 199.53 ng/ml



#27 AR-1254-2


R.T.: 7.380 min
Delta R.T.: -0.007 min
Response: 520756
Conc: 167.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#27 AR-1254-2

R.T.: 6.202 min
Delta R.T.: -0.008 min
Response: 432600
Conc: 218.42 ng/ml




#28 AR-1254-3

R.T.: 7.762 min
Delta R.T.: -0.007 min
Response: 271329
Conc: 85.29 ng/ml



#28 AR-1254-3


R.T.: 6.637 min
Delta R.T.: 0.009 min
Response: 767547
Conc: 254.83 ng/ml



#29 AR-1254-4


R.T.: 8.057 min
Delta R.T.: -0.009 min
Response: 163508
Conc: 78.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#29 AR-1254-4

R.T.: 6.860 min
Delta R.T.: -0.009 min
Response: 77037
Conc: 43.65 ng/ml




#30 AR-1254-5

R.T.: 8.487 min
Delta R.T.: -0.007 min
Response: 1638982
Conc: 713.63 ng/ml



#30 AR-1254-5


R.T.: 7.285 min
Delta R.T.: -0.010 min
Response: 1453386
Conc: 555.14 ng/ml



#31 AR-1260-1


R.T.: 7.929 min
Delta R.T.: -0.007 min
Response: 1254094
Conc: 539.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#31 AR-1260-1

R.T.: 6.755 min
Delta R.T.: -0.010 min
Response: 1093096
Conc: 532.10 ng/ml




#32 AR-1260-2

R.T.: 8.195 min
Delta R.T.: -0.007 min
Response: 1390287
Conc: 544.50 ng/ml



#32 AR-1260-2


R.T.: 6.953 min
Delta R.T.: -0.010 min
Response: 1314135
Conc: 541.86 ng/ml



#33 AR-1260-3


R.T.: 8.561 min
Delta R.T.: -0.007 min
Response: 910750
Conc: 459.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#33 AR-1260-3

R.T.: 7.106 min
Delta R.T.: -0.010 min
Response: 1210405
Conc: 535.19 ng/ml




#34 AR-1260-4

R.T.: 8.792 min
Delta R.T.: -0.007 min
Response: 1076297
Conc: 476.42 ng/ml



#34 AR-1260-4


R.T.: 7.588 min
Delta R.T.: -0.010 min
Response: 891628
Conc: 462.71 ng/ml



#35 AR-1260-5


R.T.: 9.116 min
Delta R.T.: -0.008 min
Response: 1884289
Conc: 460.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#35 AR-1260-5

R.T.: 7.836 min
Delta R.T.: -0.010 min
Response: 2057214
Conc: 465.32 ng/ml




#36 AR-1262-1

R.T.: 8.561 min
Delta R.T.: -0.006 min
Response: 910750
Conc: 307.83 ng/ml



#36 AR-1262-1


R.T.: 7.285 min
Delta R.T.: -0.009 min
Response: 1453386
Conc: 995.20 ng/ml



#37 AR-1262-2


R.T.: 9.116 min
Delta R.T.: -0.007 min
Response: 1884289
Conc: 389.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#37 AR-1262-2

R.T.: 7.836 min
Delta R.T.: -0.009 min
Response: 2057214
Conc: 415.30 ng/ml




#38 AR-1262-3

R.T.: 9.447 min
Delta R.T.: 0.014 min
Response: 1199977
Conc: 363.10 ng/ml



#38 AR-1262-3


R.T.: 8.120 min
Delta R.T.: -0.010 min
Response: 444959
Conc: 207.11 ng/ml



#39 AR-1262-4


R.T.: 9.498 min
Delta R.T.: -0.025 min
Response: 689701
Conc: 268.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#39 AR-1262-4

R.T.: 8.182 min
Delta R.T.: -0.013 min
Response: 1526081
Conc: 403.48 ng/ml




#40 AR-1262-5

R.T.: 10.175 min
Delta R.T.: -0.010 min
Response: 434558
Conc: 264.18 ng/ml



#40 AR-1262-5


R.T.: 8.680 min
Delta R.T.: -0.010 min
Response: 491234
Conc: 283.77 ng/ml



#41 AR-1268-1


R.T.: 9.447 min
Delta R.T.: 0.015 min
Response: 1199977
Conc: 206.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#41 AR-1268-1

R.T.: 8.120 min
Delta R.T.: -0.009 min
Response: 444959
Conc: 75.76 ng/ml




#42 AR-1268-2

R.T.: 9.498 min
Delta R.T.: -0.027 min
Response: 689701
Conc: 130.01 ng/ml



#42 AR-1268-2


R.T.: 8.182 min
Delta R.T.: -0.013 min
Response: 1526081
Conc: 286.52 ng/ml



#43 AR-1268-3


R.T.: 9.742 min
Delta R.T.: -0.008 min
Response: 37933
Conc: 8.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD




#43 AR-1268-3

R.T.: 8.390 min
Delta R.T.: -0.010 min
Response: 38741
Conc: 8.17 ng/ml




#44 AR-1268-4

R.T.: 10.175 min
Delta R.T.: -0.008 min
Response: 434558
Conc: 249.35 ng/ml



#44 AR-1268-4


R.T.: 8.680 min
Delta R.T.: -0.009 min
Response: 491234
Conc: 264.34 ng/ml



#45 AR-1268-5

R.T.: 10.582 min
Delta R.T.: -0.010 min
Response: 159124
Conc: 11.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB135270BSD



#45 AR-1268-5

R.T.: 8.955 min
Delta R.T.: -0.011 min
Response: 170002
Conc: 12.74 ng/ml