

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032221\
 Data File : P0076412.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Mar 2021 12:21
 Operator : DD\AJ
 Sample : M1458-03
 Misc : AR1262 MDL 25 PPB
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MDL-SOIL-03-QT1-2021

Manual Integrations
 APPROVED

Ankita
 3/23/2021 1:40:18 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 22 15:42:49 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 16 07:18:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.917	3.929	1128808	785262	19.870	18.822
2) SA Decachlor...	10.909	9.190	838344	805202	21.796	20.970m
Target Compounds						
36) L8 AR-1262-1	8.560	7.283	72663	38447	24.560	26.326
37) L8 AR-1262-2	9.115	7.835	112199	124166	23.177	25.066
38) L8 AR-1262-3	9.422	8.120	80463	56801	24.347	26.439
39) L8 AR-1262-4	9.515	8.184	64391	94709	25.049	25.040
40) L8 AR-1262-5	10.174	8.680	36682	42450	22.300	24.522

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032221\
Data File : P0076412.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Mar 2021 12:21
Operator : DD\AJ
Sample : M1458-03
Misc : AR1262 MDL 25 PPB
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
MDL-SOIL-03-QT1-2021

Manual Integrations
APPROVED

Ankita
3/23/2021 1:40:18 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 22 15:42:49 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 16 07:18:57 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

