

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032221\
 Data File : P0076430.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Mar 2021 17:28
 Operator : DD\AJ
 Sample : M1757-02
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 B1B

Manual Integrations
 APPROVED

Ankita
 3/23/2021 1:40:48 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 00:37:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031521.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 16 07:18:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.917	3.928	1489792	1050678	26.224	25.183m
2) SA Decachlor...	10.911	9.191	917372	823258	23.850	21.440

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032221\
Data File : P0076430.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Mar 2021 17:28
Operator : DD\AJ
Sample : M1757-02
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
B1B


Manual Integrations
APPROVED

Ankita
3/23/2021 1:40:48 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 00:37:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031521.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 16 07:18:57 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.917 min
Delta R.T.: -0.005 min
Response: 1489792
Conc: 26.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
B1B


Manual Integrations
APPROVED

Ankita
3/23/2021 1:40:48 PM




#1 Tetrachloro-m-xylene

R.T.: 3.928 min
Delta R.T.: -0.008 min
Response: 1050678
Conc: 25.18 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.911 min
Delta R.T.: -0.013 min
Response: 917372
Conc: 23.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.191 min
Delta R.T.: -0.013 min
Response: 823258
Conc: 21.44 ng/ml