

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032223\
 Data File : P0093415.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Mar 2023 09:16
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 00:58:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 14 05:32:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.424	3.629	134.0E6	47999160	42.802	42.558
2) SA Decachlor...	10.249	8.654	98258163	42647685	45.400	43.556
Target Compounds						
21) L5 AR-1248-1	5.600	4.714	26057445	10050648	419.173	431.883
22) L5 AR-1248-2	5.876	4.950	41727199	15384177	436.889	424.502
23) L5 AR-1248-3	6.081	4.992	44716161	16302469	429.719	434.424
24) L5 AR-1248-4	6.488	5.163	44743585	19223649	439.390	446.839
25) L5 AR-1248-5	6.527	5.557	46409504	16984126	453.398	461.809



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO032223\
Data File : PO093415.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Mar 2023 09:16
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 00:58:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 14 05:32:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.424 min
Delta R.T.: -0.006 min
Response: 134035010
Conc: 42.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.629 min
Delta R.T.: -0.006 min
Response: 47999160
Conc: 42.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.249 min
Delta R.T.: -0.014 min
Response: 98258163
Conc: 45.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.654 min
Delta R.T.: -0.014 min
Response: 42647685
Conc: 43.56 ng/ml



#21 AR-1248-1


R.T.: 5.600 min
Delta R.T.: -0.007 min
Response: 26057445
Conc: 419.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.714 min
Delta R.T.: -0.007 min
Response: 10050648
Conc: 431.88 ng/ml




#22 AR-1248-2

R.T.: 5.876 min
Delta R.T.: -0.007 min
Response: 41727199
Conc: 436.89 ng/ml



#22 AR-1248-2


R.T.: 4.950 min
Delta R.T.: -0.007 min
Response: 15384177
Conc: 424.50 ng/ml



#23 AR-1248-3


R.T.: 6.081 min
Delta R.T.: -0.007 min
Response: 44716161
Conc: 429.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.992 min
Delta R.T.: -0.008 min
Response: 16302469
Conc: 434.42 ng/ml




#24 AR-1248-4

R.T.: 6.488 min
Delta R.T.: -0.007 min
Response: 44743585
Conc: 439.39 ng/ml



#24 AR-1248-4


R.T.: 5.163 min
Delta R.T.: -0.008 min
Response: 19223649
Conc: 446.84 ng/ml



#25 AR-1248-5

R.T.: 6.527 min
Delta R.T.: -0.007 min
Response: 46409504
Conc: 453.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.557 min
Delta R.T.: -0.009 min
Response: 16984126
Conc: 461.81 ng/ml