

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032322\
 Data File : P0085576.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Mar 2022 17:07
 Operator : YP\AJ
 Sample : PB143515BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB143515BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 01:17:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 16 04:43:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.489	3.590	3258189	851048	15.934	19.864
2)	SA Decachlor...	10.387	8.604	1860056	680166	15.474	17.716
Target Compounds							
3)	L1 AR-1016-1	5.676	4.677	2116246	661480	336.278	414.493
4)	L1 AR-1016-2	5.699	4.695	3052882	951482	333.285	411.226
5)	L1 AR-1016-3	5.762	4.871	1809029	549699	332.348	439.842 #
6)	L1 AR-1016-4	5.862	4.914	1483895	469821	336.405	430.382 #
7)	L1 AR-1016-5	6.160	5.127	1396300	582492	328.837	430.649 #
8)	L2 AR-1221-1	4.697	3.803	251126	68049	113.848	131.948
9)	L2 AR-1221-2	4.788	3.889	310904	94494	190.856	225.039
10)	L2 AR-1221-3	4.865	3.965	1309129	392275	254.591	305.907
11)	L3 AR-1232-1	4.865	3.965	1309129	392275	278.174	331.095
12)	L3 AR-1232-2	5.405	4.695	1915681	951482	831.316	873.278
13)	L3 AR-1232-3	5.699	4.871	3052882	549699	723.471	904.587 #
14)	L3 AR-1232-4	5.862	4.956	1483895	549504	739.002	986.000 #
15)	L3 AR-1232-5	5.954	5.127	1137725	582492	773.015	944.714
16)	L4 AR-1242-1	5.676	4.677	2116246	661480	416.271	506.244
17)	L4 AR-1242-2	5.699	4.695	3052882	951482	411.263	514.864 #
18)	L4 AR-1242-3	5.762	4.871	1809029	549699	411.656	543.638 #
19)	L4 AR-1242-4	5.862	4.956	1483895	549504	408.936	528.648 #
20)	L4 AR-1242-5	6.609	5.479	220014	444279	61.917	359.992 #
21)	L5 AR-1248-1	5.676	4.677	2116246	661480	567.691	684.195
22)	L5 AR-1248-2	5.954	4.914	1137725	469821	214.644	320.850 #
23)	L5 AR-1248-3	6.160	4.956	1396300	549504	237.330	358.510 #
24)	L5 AR-1248-4	6.568	5.127	225549	582492	34.986	331.394 #
25)	L5 AR-1248-5	6.609	5.519	220014	69308	35.536	40.352
26)	L6 AR-1254-1	6.543	5.479	1157120	444279	174.600	169.042
27)	L6 AR-1254-2	6.763	5.628	1049892	440462	104.908	188.765 #
28)	L6 AR-1254-3	7.135	6.047	539200	757595	52.393	206.271 #
29)	L6 AR-1254-4	7.425	6.261	357450	88534	50.949	39.969
30)	L6 AR-1254-5	7.848	6.680	3405821	1389129	445.114	427.059
31)	L7 AR-1260-1	7.302	6.162	2469557	1029949	357.907	434.426
32)	L7 AR-1260-2	7.562	6.352	2824831	1210114	361.110	433.486
33)	L7 AR-1260-3	7.926	6.505	1730925	1116477	294.695	424.647 #
34)	L7 AR-1260-4	8.156	6.979	2198807	793936	318.344	354.466
35)	L7 AR-1260-5	8.486	7.223	4355326	1801340	310.007	357.564
36)	L8 AR-1262-1	7.926	6.680	1730925	1389129	191.647	809.052 #
37)	L8 AR-1262-2	8.486	6.979	4355326	793936	273.343	277.234
38)	L8 AR-1262-3	8.825	7.507	2468558	368884	238.828	168.439 #
39)	L8 AR-1262-4	8.905	7.569	512110	1289908	109.033	323.235 #
40)	L8 AR-1262-5	9.591	8.067	897506	397529	168.768	216.819 #
41)	L9 AR-1268-1	8.825	7.507	2468558	368884	127.909	55.543 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032322\
Data File : P0085576.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Mar 2022 17:07
Operator : YP\AJ
Sample : PB143515BS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB143515BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 01:17:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 04:43:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.905	7.569	512110	1289908	29.473	218.854 #
43)	L9 AR-1268-3	9.146	7.778	61072	29821	4.138	6.002 #
44)	L9 AR-1268-4	9.591	8.067	897506	397529	147.096	187.325 #
45)	L9 AR-1268-5	10.030	8.354	299162	121527	6.197	8.678 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032322\
Data File : P0085576.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Mar 2022 17:07
Operator : YP\AJ
Sample : PB143515BS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB143515BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 01:17:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 16 04:43:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.489 min
Delta R.T.: -0.003 min
Response: 3258189
Conc: 15.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#1 Tetrachloro-m-xylene

R.T.: 3.590 min
Delta R.T.: 0.000 min
Response: 851048
Conc: 19.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.387 min
Delta R.T.: -0.008 min
Response: 1860056
Conc: 15.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.604 min
Delta R.T.: -0.002 min
Response: 680166
Conc: 17.72 ng/ml



#3 AR-1016-1


R.T.: 5.676 min
Delta R.T.: -0.004 min
Response: 2116246
Conc: 336.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#3 AR-1016-1

R.T.: 4.677 min
Delta R.T.: -0.002 min
Response: 661480
Conc: 414.49 ng/ml




#4 AR-1016-2

R.T.: 5.699 min
Delta R.T.: -0.004 min
Response: 3052882
Conc: 333.28 ng/ml



#4 AR-1016-2


R.T.: 4.695 min
Delta R.T.: -0.002 min
Response: 951482
Conc: 411.23 ng/ml



#5 AR-1016-3


R.T.: 5.762 min
Delta R.T.: -0.005 min
Response: 1809029
Conc: 332.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#5 AR-1016-3

R.T.: 4.871 min
Delta R.T.: -0.002 min
Response: 549699
Conc: 439.84 ng/ml




#6 AR-1016-4

R.T.: 5.862 min
Delta R.T.: -0.004 min
Response: 1483895
Conc: 336.40 ng/ml



#6 AR-1016-4


R.T.: 4.914 min
Delta R.T.: -0.002 min
Response: 469821
Conc: 430.38 ng/ml



#7 AR-1016-5


R.T.: 6.160 min
Delta R.T.: -0.005 min
Response: 1396300
Conc: 328.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#7 AR-1016-5

R.T.: 5.127 min
Delta R.T.: -0.002 min
Response: 582492
Conc: 430.65 ng/ml




#8 AR-1221-1

R.T.: 4.697 min
Delta R.T.: -0.006 min
Response: 251126
Conc: 113.85 ng/ml



#8 AR-1221-1


R.T.: 3.803 min
Delta R.T.: -0.002 min
Response: 68049
Conc: 131.95 ng/ml



#9 AR-1221-2


R.T.: 4.788 min
Delta R.T.: -0.003 min
Response: 310904
Conc: 190.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#9 AR-1221-2

R.T.: 3.889 min
Delta R.T.: 0.000 min
Response: 94494
Conc: 225.04 ng/ml




#10 AR-1221-3

R.T.: 4.865 min
Delta R.T.: -0.003 min
Response: 1309129
Conc: 254.59 ng/ml



#10 AR-1221-3


R.T.: 3.965 min
Delta R.T.: -0.001 min
Response: 392275
Conc: 305.91 ng/ml



#11 AR-1232-1


R.T.: 4.865 min
Delta R.T.: -0.003 min
Response: 1309129
Conc: 278.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#11 AR-1232-1

R.T.: 3.965 min
Delta R.T.: 0.000 min
Response: 392275
Conc: 331.10 ng/ml




#12 AR-1232-2

R.T.: 5.405 min
Delta R.T.: -0.004 min
Response: 1915681
Conc: 831.32 ng/ml



#12 AR-1232-2


R.T.: 4.695 min
Delta R.T.: -0.002 min
Response: 951482
Conc: 873.28 ng/ml



#13 AR-1232-3


R.T.: 5.699 min
Delta R.T.: -0.004 min
Response: 3052882
Conc: 723.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#13 AR-1232-3

R.T.: 4.871 min
Delta R.T.: -0.001 min
Response: 549699
Conc: 904.59 ng/ml




#14 AR-1232-4

R.T.: 5.862 min
Delta R.T.: -0.005 min
Response: 1483895
Conc: 739.00 ng/ml



#14 AR-1232-4


R.T.: 4.956 min
Delta R.T.: -0.001 min
Response: 549504
Conc: 986.00 ng/ml



#15 AR-1232-5


R.T.: 5.954 min
Delta R.T.: -0.004 min
Response: 1137725
Conc: 773.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#15 AR-1232-5

R.T.: 5.127 min
Delta R.T.: -0.001 min
Response: 582492
Conc: 944.71 ng/ml




#16 AR-1242-1

R.T.: 5.676 min
Delta R.T.: -0.004 min
Response: 2116246
Conc: 416.27 ng/ml



#16 AR-1242-1


R.T.: 4.677 min
Delta R.T.: -0.002 min
Response: 661480
Conc: 506.24 ng/ml



#17 AR-1242-2


R.T.: 5.699 min
Delta R.T.: -0.004 min
Response: 3052882
Conc: 411.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#17 AR-1242-2

R.T.: 4.695 min
Delta R.T.: -0.002 min
Response: 951482
Conc: 514.86 ng/ml




#18 AR-1242-3

R.T.: 5.762 min
Delta R.T.: -0.004 min
Response: 1809029
Conc: 411.66 ng/ml



#18 AR-1242-3


R.T.: 4.871 min
Delta R.T.: -0.002 min
Response: 549699
Conc: 543.64 ng/ml



#19 AR-1242-4


R.T.: 5.862 min
Delta R.T.: -0.005 min
Response: 1483895
Conc: 408.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#19 AR-1242-4

R.T.: 4.956 min
Delta R.T.: -0.001 min
Response: 549504
Conc: 528.65 ng/ml




#20 AR-1242-5

R.T.: 6.609 min
Delta R.T.: -0.004 min
Response: 220014
Conc: 61.92 ng/ml



#20 AR-1242-5


R.T.: 5.479 min
Delta R.T.: -0.001 min
Response: 444279
Conc: 359.99 ng/ml



#21 AR-1248-1


R.T.: 5.676 min
Delta R.T.: -0.003 min
Response: 2116246
Conc: 567.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#21 AR-1248-1

R.T.: 4.677 min
Delta R.T.: -0.001 min
Response: 661480
Conc: 684.19 ng/ml




#22 AR-1248-2

R.T.: 5.954 min
Delta R.T.: -0.003 min
Response: 1137725
Conc: 214.64 ng/ml



#22 AR-1248-2


R.T.: 4.914 min
Delta R.T.: -0.001 min
Response: 469821
Conc: 320.85 ng/ml



#23 AR-1248-3


R.T.: 6.160 min
Delta R.T.: -0.004 min
Response: 1396300
Conc: 237.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#23 AR-1248-3

R.T.: 4.956 min
Delta R.T.: 0.000 min
Response: 549504
Conc: 358.51 ng/ml




#24 AR-1248-4

R.T.: 6.568 min
Delta R.T.: -0.005 min
Response: 225549
Conc: 34.99 ng/ml



#24 AR-1248-4


R.T.: 5.127 min
Delta R.T.: 0.000 min
Response: 582492
Conc: 331.39 ng/ml



#25 AR-1248-5


R.T.: 6.609 min
Delta R.T.: -0.003 min
Response: 220014
Conc: 35.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#25 AR-1248-5

R.T.: 5.519 min
Delta R.T.: -0.002 min
Response: 69308
Conc: 40.35 ng/ml




#26 AR-1254-1

R.T.: 6.543 min
Delta R.T.: -0.004 min
Response: 1157120
Conc: 174.60 ng/ml



#26 AR-1254-1


R.T.: 5.479 min
Delta R.T.: -0.001 min
Response: 444279
Conc: 169.04 ng/ml



#27 AR-1254-2


R.T.: 6.763 min
Delta R.T.: -0.005 min
Response: 1049892
Conc: 104.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#27 AR-1254-2

R.T.: 5.628 min
Delta R.T.: 0.000 min
Response: 440462
Conc: 188.77 ng/ml




#28 AR-1254-3

R.T.: 7.135 min
Delta R.T.: -0.003 min
Response: 539200
Conc: 52.39 ng/ml



#28 AR-1254-3


R.T.: 6.047 min
Delta R.T.: 0.015 min
Response: 757595
Conc: 206.27 ng/ml



#29 AR-1254-4


R.T.: 7.425 min
Delta R.T.: -0.003 min
Response: 357450
Conc: 50.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#29 AR-1254-4

R.T.: 6.261 min
Delta R.T.: 0.000 min
Response: 88534
Conc: 39.97 ng/ml




#30 AR-1254-5

R.T.: 7.848 min
Delta R.T.: -0.004 min
Response: 3405821
Conc: 445.11 ng/ml



#30 AR-1254-5


R.T.: 6.680 min
Delta R.T.: 0.000 min
Response: 1389129
Conc: 427.06 ng/ml



#31 AR-1260-1


R.T.: 7.302 min
Delta R.T.: -0.005 min
Response: 2469557
Conc: 357.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#31 AR-1260-1

R.T.: 6.162 min
Delta R.T.: -0.002 min
Response: 1029949
Conc: 434.43 ng/ml




#32 AR-1260-2

R.T.: 7.562 min
Delta R.T.: -0.004 min
Response: 2824831
Conc: 361.11 ng/ml



#32 AR-1260-2


R.T.: 6.352 min
Delta R.T.: -0.003 min
Response: 1210114
Conc: 433.49 ng/ml



#33 AR-1260-3


R.T.: 7.926 min
Delta R.T.: -0.004 min
Response: 1730925
Conc: 294.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#33 AR-1260-3

R.T.: 6.505 min
Delta R.T.: -0.003 min
Response: 1116477
Conc: 424.65 ng/ml




#34 AR-1260-4

R.T.: 8.156 min
Delta R.T.: -0.004 min
Response: 2198807
Conc: 318.34 ng/ml



#34 AR-1260-4


R.T.: 6.979 min
Delta R.T.: -0.002 min
Response: 793936
Conc: 354.47 ng/ml



#35 AR-1260-5


R.T.: 8.486 min
Delta R.T.: -0.006 min
Response: 4355326
Conc: 310.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#35 AR-1260-5

R.T.: 7.223 min
Delta R.T.: -0.002 min
Response: 1801340
Conc: 357.56 ng/ml




#36 AR-1262-1

R.T.: 7.926 min
Delta R.T.: -0.004 min
Response: 1730925
Conc: 191.65 ng/ml



#36 AR-1262-1


R.T.: 6.680 min
Delta R.T.: 0.000 min
Response: 1389129
Conc: 809.05 ng/ml



#37 AR-1262-2


R.T.: 8.486 min
Delta R.T.: -0.005 min
Response: 4355326
Conc: 273.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#37 AR-1262-2

R.T.: 6.979 min
Delta R.T.: 0.000 min
Response: 793936
Conc: 277.23 ng/ml




#38 AR-1262-3

R.T.: 8.825 min
Delta R.T.: 0.007 min
Response: 2468558
Conc: 238.83 ng/ml



#38 AR-1262-3


R.T.: 7.507 min
Delta R.T.: 0.000 min
Response: 368884
Conc: 168.44 ng/ml



#39 AR-1262-4


R.T.: 8.905 min
Delta R.T.: -0.006 min
Response: 512110
Conc: 109.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#39 AR-1262-4

R.T.: 7.569 min
Delta R.T.: -0.002 min
Response: 1289908
Conc: 323.24 ng/ml




#40 AR-1262-5

R.T.: 9.591 min
Delta R.T.: -0.006 min
Response: 897506
Conc: 168.77 ng/ml



#40 AR-1262-5


R.T.: 8.067 min
Delta R.T.: 0.000 min
Response: 397529
Conc: 216.82 ng/ml



#41 AR-1268-1


R.T.: 8.825 min
Delta R.T.: 0.011 min
Response: 2468558
Conc: 127.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#41 AR-1268-1

R.T.: 7.507 min
Delta R.T.: 0.000 min
Response: 368884
Conc: 55.54 ng/ml




#42 AR-1268-2

R.T.: 8.905 min
Delta R.T.: -0.007 min
Response: 512110
Conc: 29.47 ng/ml



#42 AR-1268-2


R.T.: 7.569 min
Delta R.T.: -0.003 min
Response: 1289908
Conc: 218.85 ng/ml



#43 AR-1268-3


R.T.: 9.146 min
Delta R.T.: -0.006 min
Response: 61072
Conc: 4.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS




#43 AR-1268-3

R.T.: 7.778 min
Delta R.T.: -0.001 min
Response: 29821
Conc: 6.00 ng/ml




#44 AR-1268-4

R.T.: 9.591 min
Delta R.T.: -0.005 min
Response: 897506
Conc: 147.10 ng/ml



#44 AR-1268-4


R.T.: 8.067 min
Delta R.T.: 0.000 min
Response: 397529
Conc: 187.32 ng/ml



#45 AR-1268-5

R.T.: 10.030 min
Delta R.T.: -0.003 min
Response: 299162
Conc: 6.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB143515BS



#45 AR-1268-5

R.T.: 8.354 min
Delta R.T.: 0.000 min
Response: 121527
Conc: 8.68 ng/ml