

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032421\
 Data File : P0076521.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Mar 2021 11:16
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Ankita
 3/25/2021 4:44:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 25 08:48:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 24 05:38:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.922	3.928	4375832	2514558	49.964	45.841
2) SA Decachlor...	10.896	9.187	3338558	2235008	51.288	47.431
Target Compounds						
3) L1 AR-1016-1	6.243	5.177	1345282	739911	497.789	449.240
4) L1 AR-1016-2	6.267	5.197	2144271	1367210	521.608	471.127
5) L1 AR-1016-3	6.333	5.386	1257367	622993	521.113	466.590
6) L1 AR-1016-4	6.440	5.436	1008013	585103	520.079	521.255
7) L1 AR-1016-5	6.753	5.664	1091561	628071	560.434	456.118
31) L7 AR-1260-1	7.925	6.753	1855868	1270543	540.065m	495.749m
32) L7 AR-1260-2	8.192	6.951	2176410	1528845	536.778m	492.970m
33) L7 AR-1260-3	8.557	7.104	1725471	1431438	571.133	487.256m
34) L7 AR-1260-4	8.787	7.585	1898811	1221057	544.881	512.333
35) L7 AR-1260-5	9.108	7.833	3582894	2817042	530.183	496.541

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032421\
Data File : P0076521.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Mar 2021 11:16
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Ankita
3/25/2021 4:44:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 25 08:48:35 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 24 05:38:14 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

