

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032421\
 Data File : P0076493.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Mar 2021 2:54
 Operator : DD\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 03:17:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 24 03:09:33 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.914	3.928	2370994	1420905	24.432	25.426
2) SA Decachlor...	10.889	9.187	3036999	2084508	25.161	25.551
Target Compounds						
41) L9 AR-1268-1	9.409	8.117	2581810	1961360	248.584	254.212
42) L9 AR-1268-2	9.505	8.182	2309635	1778578	247.768	252.809
43) L9 AR-1268-3	9.726	8.388	2063304	1536880	250.010	254.839
44) L9 AR-1268-4	10.156	8.678	805944	617990	250.022	254.877
45) L9 AR-1268-5	10.561	8.951	6601470	4509187	247.650	249.013



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032421\
Data File : P0076493.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Mar 2021 2:54
Operator : DD\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 03:17:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 24 03:09:33 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.914 min
Delta R.T.: 0.000 min
Response: 2370994
Conc: 24.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.928 min
Delta R.T.: 0.000 min
Response: 1420905
Conc: 25.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.889 min
Delta R.T.: 0.000 min
Response: 3036999
Conc: 25.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.187 min
Delta R.T.: 0.000 min
Response: 2084508
Conc: 25.55 ng/ml



#41 AR-1268-1


R.T.: 9.409 min
Delta R.T.: 0.000 min
Response: 2581810
Conc: 248.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 8.117 min
Delta R.T.: 0.000 min
Response: 1961360
Conc: 254.21 ng/ml




#42 AR-1268-2

R.T.: 9.505 min
Delta R.T.: 0.000 min
Response: 2309635
Conc: 247.77 ng/ml



#42 AR-1268-2


R.T.: 8.182 min
Delta R.T.: 0.000 min
Response: 1778578
Conc: 252.81 ng/ml



#43 AR-1268-3


R.T.: 9.726 min
Delta R.T.: 0.000 min
Response: 2063304
Conc: 250.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 8.388 min
Delta R.T.: 0.000 min
Response: 1536880
Conc: 254.84 ng/ml




#44 AR-1268-4

R.T.: 10.156 min
Delta R.T.: 0.000 min
Response: 805944
Conc: 250.02 ng/ml



#44 AR-1268-4


R.T.: 8.678 min
Delta R.T.: 0.000 min
Response: 617990
Conc: 254.88 ng/ml



#45 AR-1268-5

R.T.: 10.561 min
Delta R.T.: 0.000 min
Response: 6601470
Conc: 247.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.951 min
Delta R.T.: 0.000 min
Response: 4509187
Conc: 249.01 ng/ml