

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032421\
 Data File : P0076497.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Mar 2021 4:02
 Operator : DD\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO032421AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 04:17:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 24 04:16:18 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.914	3.928	5063446	2914027	54.082	51.882
2) SA Decachlor...	10.887	9.187	3559194	2436061	53.025	51.404
Target Compounds						
21) L5 AR-1248-1	6.236	5.177	897723	502819	542.335	501.698
22) L5 AR-1248-2	6.529	5.437	1316716	854904	539.406	525.062
23) L5 AR-1248-3	6.747	5.481	1564833	861299	546.209	521.795
24) L5 AR-1248-4	7.175	5.664	1690926	1013565	541.521	521.346
25) L5 AR-1248-5	7.215	6.086	1725749	989498	554.217	534.010



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO032421\
Data File : PO076497.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Mar 2021 4:02
Operator : DD\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO032421AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 04:17:20 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO032421.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 24 04:16:18 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.914 min
Delta R.T.: 0.000 min
Response: 5063446
Conc: 54.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO032421AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.928 min
Delta R.T.: 0.000 min
Response: 2914027
Conc: 51.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.887 min
Delta R.T.: -0.001 min
Response: 3559194
Conc: 53.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.187 min
Delta R.T.: 0.000 min
Response: 2436061
Conc: 51.40 ng/ml



#21 AR-1248-1


R.T.: 6.236 min
Delta R.T.: 0.000 min
Response: 897723
Conc: 542.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO032421AR1248




#21 AR-1248-1

R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 502819
Conc: 501.70 ng/ml




#22 AR-1248-2

R.T.: 6.529 min
Delta R.T.: 0.000 min
Response: 1316716
Conc: 539.41 ng/ml



#22 AR-1248-2


R.T.: 5.437 min
Delta R.T.: 0.000 min
Response: 854904
Conc: 525.06 ng/ml



#23 AR-1248-3


R.T.: 6.747 min
Delta R.T.: 0.000 min
Response: 1564833
Conc: 546.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO032421AR1248




#23 AR-1248-3

R.T.: 5.481 min
Delta R.T.: 0.000 min
Response: 861299
Conc: 521.79 ng/ml




#24 AR-1248-4

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 1690926
Conc: 541.52 ng/ml



#24 AR-1248-4


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 1013565
Conc: 521.35 ng/ml



#25 AR-1248-5

R.T.: 7.215 min
Delta R.T.: 0.000 min
Response: 1725749
Conc: 554.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO032421AR1248



#25 AR-1248-5

R.T.: 6.086 min
Delta R.T.: 0.000 min
Response: 989498
Conc: 534.01 ng/ml