

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032423\
 Data File : P0093515.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Mar 2023 15:26
 Operator : YP/AJ
 Sample : 02068-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CONCRETE-TCLP

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 21:24:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 14 05:32:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.426	3.632	57375126	21726235	18.322	19.263
2) SA Decachlor...	10.256	8.658	44302056	19497696	20.470	19.913
Target Compounds						
31) L7 AR-1260-1	7.221	6.203	110.4E6	48914877	852.007	804.541
32) L7 AR-1260-2	7.479	6.392	179.5E6	65953930	1215.651	912.267
33) L7 AR-1260-3	7.843	6.546	91894790	54350603	816.982	812.797
34) L7 AR-1260-4	8.070	7.019	107.6E6	39719242	838.746	721.754
35) L7 AR-1260-5	8.397	7.263	218.3E6	100.2E6	977.191	851.324



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO032423\
Data File : PO093515.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Mar 2023 15:26
Operator : YP/AJ
Sample : 02068-01
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CONCRETE-TCLP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 21:24:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 14 05:32:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.426 min
Delta R.T.: -0.004 min
Response: 57375126
Conc: 18.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
CONCRETE-TCLP




#1 Tetrachloro-m-xylene

R.T.: 3.632 min
Delta R.T.: -0.003 min
Response: 21726235
Conc: 19.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.256 min
Delta R.T.: -0.007 min
Response: 44302056
Conc: 20.47 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.658 min
Delta R.T.: -0.010 min
Response: 19497696
Conc: 19.91 ng/ml



#31 AR-1260-1


R.T.: 7.221 min
Delta R.T.: -0.005 min
Response: 110379695
Conc: 852.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
CONCRETE-TCLP




#31 AR-1260-1

R.T.: 6.203 min
Delta R.T.: -0.007 min
Response: 48914877
Conc: 804.54 ng/ml




#32 AR-1260-2

R.T.: 7.479 min
Delta R.T.: -0.006 min
Response: 179460429
Conc: 1215.65 ng/ml



#32 AR-1260-2


R.T.: 6.392 min
Delta R.T.: -0.006 min
Response: 65953930
Conc: 912.27 ng/ml



#33 AR-1260-3


R.T.: 7.843 min
Delta R.T.: -0.005 min
Response: 91894790
Conc: 816.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
CONCRETE-TCLP




#33 AR-1260-3

R.T.: 6.546 min
Delta R.T.: -0.007 min
Response: 54350603
Conc: 812.80 ng/ml




#34 AR-1260-4

R.T.: 8.070 min
Delta R.T.: -0.005 min
Response: 107559007
Conc: 838.75 ng/ml



#34 AR-1260-4


R.T.: 7.019 min
Delta R.T.: -0.008 min
Response: 39719242
Conc: 721.75 ng/ml



#35 AR-1260-5

R.T.: 8.397 min
Delta R.T.: -0.005 min
Response: 218338850
Conc: 977.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
CONCRETE-TCLP



#35 AR-1260-5

R.T.: 7.263 min
Delta R.T.: -0.007 min
Response: 100227006
Conc: 851.32 ng/ml