

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032423\
 Data File : P0093532.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Mar 2023 01:38
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 21:34:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 14 05:32:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.424	3.630	153.0E6	53579912	48.844	47.506
2) SA Decachlor...	10.251	8.654	114.2E6	49147023	52.780	50.194
Target Compounds						
21) L5 AR-1248-1	5.601	4.713	28156596	10601465	452.941	455.552
22) L5 AR-1248-2	5.877	4.950	44528937	16556434	466.224	456.848
23) L5 AR-1248-3	6.082	4.992	48196036	17270995	463.160	460.233
24) L5 AR-1248-4	6.489	5.164	47848914	19787459	469.885	459.944
25) L5 AR-1248-5	6.528	5.557	49360358	17494511	482.226	475.686



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032423\
Data File : P0093532.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Mar 2023 01:38
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 21:34:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 14 05:32:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.424 min
Delta R.T.: -0.006 min
Response: 152953842
Conc: 48.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.630 min
Delta R.T.: -0.005 min
Response: 53579912
Conc: 47.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.251 min
Delta R.T.: -0.013 min
Response: 114231851
Conc: 52.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.654 min
Delta R.T.: -0.014 min
Response: 49147023
Conc: 50.19 ng/ml



#21 AR-1248-1


R.T.: 5.601 min
Delta R.T.: -0.005 min
Response: 28156596
Conc: 452.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.713 min
Delta R.T.: -0.008 min
Response: 10601465
Conc: 455.55 ng/ml




#22 AR-1248-2

R.T.: 5.877 min
Delta R.T.: -0.006 min
Response: 44528937
Conc: 466.22 ng/ml



#22 AR-1248-2


R.T.: 4.950 min
Delta R.T.: -0.007 min
Response: 16556434
Conc: 456.85 ng/ml



#23 AR-1248-3


R.T.: 6.082 min
Delta R.T.: -0.006 min
Response: 48196036
Conc: 463.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.992 min
Delta R.T.: -0.008 min
Response: 17270995
Conc: 460.23 ng/ml




#24 AR-1248-4

R.T.: 6.489 min
Delta R.T.: -0.006 min
Response: 47848914
Conc: 469.88 ng/ml



#24 AR-1248-4


R.T.: 5.164 min
Delta R.T.: -0.008 min
Response: 19787459
Conc: 459.94 ng/ml



#25 AR-1248-5

R.T.: 6.528 min
Delta R.T.: -0.006 min
Response: 49360358
Conc: 482.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.557 min
Delta R.T.: -0.008 min
Response: 17494511
Conc: 475.69 ng/ml