

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032521\
 Data File : P0076530.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Mar 2021 13:52
 Operator : DD\AJ
 Sample : PB135320BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 25 02:45:40 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 24 05:38:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.914	3.928	2014346	1112960	23.000	20.290
2)	SA Decachlor...	10.888	9.184	1732738	1165867	26.619	24.742
Target Compounds							
3)	L1 AR-1016-1	6.235	5.177	1596112	857953	590.603	520.910
4)	L1 AR-1016-2	6.260	5.196	2483993	1581420	604.247	544.942
5)	L1 AR-1016-3	6.325	5.386	1483714	720621	614.922	539.708
6)	L1 AR-1016-4	6.432	5.436	1183242	654206	610.487	582.816
7)	L1 AR-1016-5	6.746	5.663	1197334	787223	614.740	571.697
8)	L2 AR-1221-1	5.161	4.185	162287	84280	184.447	145.102
9)	L2 AR-1221-2	5.264	4.284	194757	107235	287.589	248.858
10)	L2 AR-1221-3	5.352	4.372	874237	500027	404.018	372.640
11)	L3 AR-1232-1	5.352	4.372	874237	500027	494.221	457.723
12)	L3 AR-1232-2	5.940	5.196	1248548	1581420	1407.693	1323.094
13)	L3 AR-1232-3	6.260	5.386	2483993	720621	1470.997	1319.761
14)	L3 AR-1232-4	6.432	5.481	1183242	754555	1511.208	1469.719
15)	L3 AR-1232-5	6.529	5.663	976884	787223	1772.074	1443.107
16)	L4 AR-1242-1	6.235	5.177	1596112	857953	713.965	666.126
17)	L4 AR-1242-2	6.260	5.196	2483993	1581420	731.096	674.815
18)	L4 AR-1242-3	6.325	5.386	1483714	720621	734.185	665.510
19)	L4 AR-1242-4	6.432	5.481	1183242	754555	734.671	681.041
20)	L4 AR-1242-5	7.215	6.040	173184	655340	99.151	494.066 #
21)	L5 AR-1248-1	6.235	5.177	1596112	857953	964.249	856.039
22)	L5 AR-1248-2	6.529	5.436	976884	654206	400.190	401.797
23)	L5 AR-1248-3	6.746	5.481	1197334	754555	417.932	457.127
24)	L5 AR-1248-4	7.175	5.663	185781	787223	59.497	404.922 #
25)	L5 AR-1248-5	7.215	6.086	173184	93186	55.617	50.290
26)	L6 AR-1254-1	7.142	6.040	921565	655340	277.456	221.255
27)	L6 AR-1254-2	7.372	6.199	1011152	624542	196.458	240.380
28)	L6 AR-1254-3	7.754	6.634f	515590	1102995	96.929	275.391 #
29)	L6 AR-1254-4	8.047	6.856	315891	109496	87.584	46.211 #
30)	L6 AR-1254-5	8.477	7.281	3173981	2153866	809.728	611.101
31)	L7 AR-1260-1	7.920	6.752	2327306	1578940	677.256	616.081
32)	L7 AR-1260-2	8.186	6.949	2666880	1911303	657.745	616.292
33)	L7 AR-1260-3	8.550	7.102	1726800	1767894	571.572	601.784
34)	L7 AR-1260-4	8.781	7.583	2048893	1292512	587.948	542.315
35)	L7 AR-1260-5	9.103	7.831	3842737	3031910	568.633	534.414
36)	L8 AR-1262-1	8.550	7.281	1726800	2153866	333.751	1095.285 #
37)	L8 AR-1262-2	9.103	7.831	3842737	3031910	445.900	450.982
38)	L8 AR-1262-3	9.431f	8.115	2411624	606495	425.377	225.177 #
39)	L8 AR-1262-4	9.483f	8.178	1356446	2188170	315.086	439.699 #
40)	L8 AR-1262-5	10.155	8.675	899920	670025	309.722	308.382
41)	L9 AR-1268-1	9.431f	8.115	2411624	606495	232.431	77.800 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032521\
 Data File : P0076530.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Mar 2021 13:52
 Operator : DD\AJ
 Sample : PB135320BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 25 02:45:40 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 24 05:38:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.483f	8.178	1356446	2188170	145.767	308.496 #
43) L9	AR-1268-3	9.724	8.386	69081	55143	8.316	9.093
44) L9	AR-1268-4	10.155	8.675	899920	670025	276.931	277.458
45) L9	AR-1268-5	10.561	8.949	325838	228999	12.311	12.656



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032521\
Data File : P0076530.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Mar 2021 13:52
Operator : DD\AJ
Sample : PB135320BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 25 02:45:40 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 24 05:38:14 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.914 min
Delta R.T.: 0.002 min
Response: 2014346
Conc: 23.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.928 min
Delta R.T.: 0.001 min
Response: 1112960
Conc: 20.29 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.888 min
Delta R.T.: 0.010 min
Response: 1732738
Conc: 26.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.184 min
Delta R.T.: -0.002 min
Response: 1165867
Conc: 24.74 ng/ml



#3 AR-1016-1


R.T.: 6.235 min
Delta R.T.: 0.003 min
Response: 1596112
Conc: 590.60 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 5.177 min
Delta R.T.: 0.001 min
Response: 857953
Conc: 520.91 ng/ml




#4 AR-1016-2

R.T.: 6.260 min
Delta R.T.: 0.004 min
Response: 2483993
Conc: 604.25 ng/ml



#4 AR-1016-2


R.T.: 5.196 min
Delta R.T.: 0.000 min
Response: 1581420
Conc: 544.94 ng/ml



#5 AR-1016-3


R.T.: 6.325 min
Delta R.T.: 0.003 min
Response: 1483714
Conc: 614.92 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 5.386 min
Delta R.T.: 0.002 min
Response: 720621
Conc: 539.71 ng/ml




#6 AR-1016-4

R.T.: 6.432 min
Delta R.T.: 0.003 min
Response: 1183242
Conc: 610.49 ng/ml



#6 AR-1016-4


R.T.: 5.436 min
Delta R.T.: 0.000 min
Response: 654206
Conc: 582.82 ng/ml



#7 AR-1016-5


R.T.: 6.746 min
Delta R.T.: 0.003 min
Response: 1197334
Conc: 614.74 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.663 min
Delta R.T.: 0.000 min
Response: 787223
Conc: 571.70 ng/ml




#8 AR-1221-1

R.T.: 5.161 min
Delta R.T.: -0.006 min
Response: 162287
Conc: 184.45 ng/ml



#8 AR-1221-1


R.T.: 4.185 min
Delta R.T.: -0.002 min
Response: 84280
Conc: 145.10 ng/ml



#9 AR-1221-2


R.T.: 5.264 min
Delta R.T.: -0.002 min
Response: 194757
Conc: 287.59 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 4.284 min
Delta R.T.: 0.000 min
Response: 107235
Conc: 248.86 ng/ml




#10 AR-1221-3

R.T.: 5.352 min
Delta R.T.: 0.000 min
Response: 874237
Conc: 404.02 ng/ml



#10 AR-1221-3


R.T.: 4.372 min
Delta R.T.: -0.001 min
Response: 500027
Conc: 372.64 ng/ml



#11 AR-1232-1


R.T.: 5.352 min
Delta R.T.: 0.000 min
Response: 874237
Conc: 494.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.372 min
Delta R.T.: 0.000 min
Response: 500027
Conc: 457.72 ng/ml




#12 AR-1232-2

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 1248548
Conc: 1407.69 ng/ml



#12 AR-1232-2


R.T.: 5.196 min
Delta R.T.: 0.000 min
Response: 1581420
Conc: 1323.09 ng/ml



#13 AR-1232-3


R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 2483993
Conc: 1471.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 5.386 min
Delta R.T.: 0.000 min
Response: 720621
Conc: 1319.76 ng/ml




#14 AR-1232-4

R.T.: 6.432 min
Delta R.T.: 0.000 min
Response: 1183242
Conc: 1511.21 ng/ml



#14 AR-1232-4


R.T.: 5.481 min
Delta R.T.: -0.001 min
Response: 754555
Conc: 1469.72 ng/ml



#15 AR-1232-5


R.T.: 6.529 min
Delta R.T.: -0.001 min
Response: 976884
Conc: 1772.07 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.663 min
Delta R.T.: -0.001 min
Response: 787223
Conc: 1443.11 ng/ml




#16 AR-1242-1

R.T.: 6.235 min
Delta R.T.: 0.000 min
Response: 1596112
Conc: 713.96 ng/ml



#16 AR-1242-1


R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 857953
Conc: 666.13 ng/ml



#17 AR-1242-2


R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 2483993
Conc: 731.10 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 5.196 min
Delta R.T.: 0.000 min
Response: 1581420
Conc: 674.82 ng/ml




#18 AR-1242-3

R.T.: 6.325 min
Delta R.T.: 0.000 min
Response: 1483714
Conc: 734.18 ng/ml



#18 AR-1242-3


R.T.: 5.386 min
Delta R.T.: 0.000 min
Response: 720621
Conc: 665.51 ng/ml



#19 AR-1242-4


R.T.: 6.432 min
Delta R.T.: 0.000 min
Response: 1183242
Conc: 734.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.481 min
Delta R.T.: 0.000 min
Response: 754555
Conc: 681.04 ng/ml




#20 AR-1242-5

R.T.: 7.215 min
Delta R.T.: 0.000 min
Response: 173184
Conc: 99.15 ng/ml



#20 AR-1242-5


R.T.: 6.040 min
Delta R.T.: -0.002 min
Response: 655340
Conc: 494.07 ng/ml



#21 AR-1248-1


R.T.: 6.235 min
Delta R.T.: 0.000 min
Response: 1596112
Conc: 964.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 857953
Conc: 856.04 ng/ml




#22 AR-1248-2

R.T.: 6.529 min
Delta R.T.: 0.000 min
Response: 976884
Conc: 400.19 ng/ml



#22 AR-1248-2


R.T.: 5.436 min
Delta R.T.: 0.000 min
Response: 654206
Conc: 401.80 ng/ml



#23 AR-1248-3


R.T.: 6.746 min
Delta R.T.: 0.000 min
Response: 1197334
Conc: 417.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.481 min
Delta R.T.: 0.000 min
Response: 754555
Conc: 457.13 ng/ml




#24 AR-1248-4

R.T.: 7.175 min
Delta R.T.: 0.000 min
Response: 185781
Conc: 59.50 ng/ml



#24 AR-1248-4


R.T.: 5.663 min
Delta R.T.: -0.001 min
Response: 787223
Conc: 404.92 ng/ml



#25 AR-1248-5


R.T.: 7.215 min
Delta R.T.: 0.000 min
Response: 173184
Conc: 55.62 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 6.086 min
Delta R.T.: 0.000 min
Response: 93186
Conc: 50.29 ng/ml




#26 AR-1254-1

R.T.: 7.142 min
Delta R.T.: -0.003 min
Response: 921565
Conc: 277.46 ng/ml



#26 AR-1254-1


R.T.: 6.040 min
Delta R.T.: -0.003 min
Response: 655340
Conc: 221.26 ng/ml



#27 AR-1254-2


R.T.: 7.372 min
Delta R.T.: -0.002 min
Response: 1011152
Conc: 196.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 6.199 min
Delta R.T.: -0.002 min
Response: 624542
Conc: 240.38 ng/ml




#28 AR-1254-3

R.T.: 7.754 min
Delta R.T.: -0.001 min
Response: 515590
Conc: 96.93 ng/ml



#28 AR-1254-3


R.T.: 6.634 min
Delta R.T.: 0.015 min
Response: 1102995
Conc: 275.39 ng/ml



#29 AR-1254-4


R.T.: 8.047 min
Delta R.T.: -0.004 min
Response: 315891
Conc: 87.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.856 min
Delta R.T.: -0.003 min
Response: 109496
Conc: 46.21 ng/ml




#30 AR-1254-5

R.T.: 8.477 min
Delta R.T.: -0.002 min
Response: 3173981
Conc: 809.73 ng/ml



#30 AR-1254-5


R.T.: 7.281 min
Delta R.T.: -0.004 min
Response: 2153866
Conc: 611.10 ng/ml



#31 AR-1260-1


R.T.: 7.920 min
Delta R.T.: 0.004 min
Response: 2327306
Conc: 677.26 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.752 min
Delta R.T.: 0.000 min
Response: 1578940
Conc: 616.08 ng/ml




#32 AR-1260-2

R.T.: 8.186 min
Delta R.T.: 0.004 min
Response: 2666880
Conc: 657.75 ng/ml



#32 AR-1260-2


R.T.: 6.949 min
Delta R.T.: 0.000 min
Response: 1911303
Conc: 616.29 ng/ml



#33 AR-1260-3


R.T.: 8.550 min
Delta R.T.: 0.004 min
Response: 1726800
Conc: 571.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 7.102 min
Delta R.T.: 0.000 min
Response: 1767894
Conc: 601.78 ng/ml




#34 AR-1260-4

R.T.: 8.781 min
Delta R.T.: 0.005 min
Response: 2048893
Conc: 587.95 ng/ml



#34 AR-1260-4


R.T.: 7.583 min
Delta R.T.: 0.000 min
Response: 1292512
Conc: 542.31 ng/ml



#35 AR-1260-5


R.T.: 9.103 min
Delta R.T.: 0.005 min
Response: 3842737
Conc: 568.63 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.831 min
Delta R.T.: 0.000 min
Response: 3031910
Conc: 534.41 ng/ml




#36 AR-1262-1

R.T.: 8.550 min
Delta R.T.: 0.000 min
Response: 1726800
Conc: 333.75 ng/ml



#36 AR-1262-1


R.T.: 7.281 min
Delta R.T.: 0.000 min
Response: 2153866
Conc: 1095.28 ng/ml



#37 AR-1262-2


R.T.: 9.103 min
Delta R.T.: 0.000 min
Response: 3842737
Conc: 445.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.831 min
Delta R.T.: 0.000 min
Response: 3031910
Conc: 450.98 ng/ml




#38 AR-1262-3

R.T.: 9.431 min
Delta R.T.: 0.021 min
Response: 2411624
Conc: 425.38 ng/ml



#38 AR-1262-3


R.T.: 8.115 min
Delta R.T.: -0.002 min
Response: 606495
Conc: 225.18 ng/ml



#39 AR-1262-4


R.T.: 9.483 min
Delta R.T.: -0.016 min
Response: 1356446
Conc: 315.09 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 8.178 min
Delta R.T.: -0.003 min
Response: 2188170
Conc: 439.70 ng/ml




#40 AR-1262-5

R.T.: 10.155 min
Delta R.T.: 0.000 min
Response: 899920
Conc: 309.72 ng/ml



#40 AR-1262-5


R.T.: 8.675 min
Delta R.T.: -0.001 min
Response: 670025
Conc: 308.38 ng/ml



#41 AR-1268-1


R.T.: 9.431 min
Delta R.T.: 0.022 min
Response: 2411624
Conc: 232.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 8.115 min
Delta R.T.: -0.001 min
Response: 606495
Conc: 77.80 ng/ml




#42 AR-1268-2

R.T.: 9.483 min
Delta R.T.: -0.021 min
Response: 1356446
Conc: 145.77 ng/ml



#42 AR-1268-2


R.T.: 8.178 min
Delta R.T.: -0.004 min
Response: 2188170
Conc: 308.50 ng/ml



#43 AR-1268-3


R.T.: 9.724 min
Delta R.T.: 0.000 min
Response: 69081
Conc: 8.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 8.386 min
Delta R.T.: -0.001 min
Response: 55143
Conc: 9.09 ng/ml




#44 AR-1268-4

R.T.: 10.155 min
Delta R.T.: 0.000 min
Response: 899920
Conc: 276.93 ng/ml



#44 AR-1268-4


R.T.: 8.675 min
Delta R.T.: 0.000 min
Response: 670025
Conc: 277.46 ng/ml



#45 AR-1268-5

R.T.: 10.561 min
Delta R.T.: 0.000 min
Response: 325838
Conc: 12.31 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.949 min
Delta R.T.: -0.001 min
Response: 228999
Conc: 12.66 ng/ml