

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032822\
 Data File : P0085690.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Mar 2022 22:34
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO032822AR1254

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/29/2022
 Supervised By :mohammad ahmed 03/30/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 29 05:19:48 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 29 05:13:57 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.459	3.560	12015293	2609043	56.123	52.607
2) SA Decachlor...	10.352	8.577	7678553	1853773	54.685	50.725
Target Compounds						
26) L6 AR-1254-1	6.518	5.451	8348777	1453729	412.725m	524.068 #
27) L6 AR-1254-2	6.738	5.601	5819834	1384873	536.461	541.604
28) L6 AR-1254-3	7.109	6.004	7009752	1920245	516.346	506.224
29) L6 AR-1254-4	7.398	6.234	4545779	1213092	525.588	499.904
30) L6 AR-1254-5	7.824	6.654	4440824	1706222	525.994	530.893

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032822\
Data File : P0085690.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Mar 2022 22:34
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO032822AR1254

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/29/2022
Supervised By :mohammad ahmed 03/30/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 29 05:19:48 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 29 05:13:57 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

