

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032823\
 Data File : P0093634.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Mar 2023 22:07
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/29/2023
 Supervised By :Ankita Jodhani 03/29/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 29 01:04:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 14 05:32:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.422	3.628	169.3E6	59214612	54.065	52.502
2) SA Decachlor...	10.248	8.651	123.3E6	52347735	56.993	53.463
Target Compounds						
3) L1 AR-1016-1	5.599	4.712	50856683	18445659	532.241	515.303
4) L1 AR-1016-2	5.621	4.730	76558989	27173058	536.158	523.604
5) L1 AR-1016-3	5.683	4.906	49068462	14679166	540.796	530.598
6) L1 AR-1016-4	5.783	4.949	37442158	12711467	552.642	511.244
7) L1 AR-1016-5	6.080	5.161	40431514	16463914	540.007	520.206
31) L7 AR-1260-1	7.215	6.198	73437121	32376648	566.852m	532.524
32) L7 AR-1260-2	7.475	6.387	81401043	37950894	551.404	524.932
33) L7 AR-1260-3	7.838	6.540	63664051	34753562	565.999	519.729
34) L7 AR-1260-4	8.065	7.014	69721441	29114819	543.688	529.057
35) L7 AR-1260-5	8.392	7.259	128.3E6	62948948	574.190	534.686

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO032823\
Data File : PO093634.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Mar 2023 22:07
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/29/2023
Supervised By :Ankita Jodhani 03/29/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 29 01:04:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 14 05:32:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

