

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032823\
 Data File : P0093618.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Mar 2023 16:10
 Operator : YP/AJ
 Sample : 02109-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 B-30-SB01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 17:46:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 14 05:32:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.423	3.629	66265880	23330344	21.161	20.685
2) SA Decachlor...	10.252	8.653	40060691	17067234	18.510	17.431
Target Compounds						
31) L7 AR-1260-1	7.222	6.199	2703289	977441	20.866	16.077
32) L7 AR-1260-2	7.476	6.388	4449330	1315065	30.139	18.190 #
33) L7 AR-1260-3	7.840	6.541	2391727	946063	21.263	14.148 #
34) L7 AR-1260-4	8.062	7.015	2989813	735582	23.315	13.367 #
35) L7 AR-1260-5	8.395	7.259	3294187	1451616	14.743	12.330



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032823\
Data File : P0093618.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Mar 2023 16:10
Operator : YP/AJ
Sample : 02109-01
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
B-30-SB01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 17:46:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 14 05:32:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.423 min
Delta R.T.: -0.007 min
Response: 66265880
Conc: 21.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
B-30-SB01




#1 Tetrachloro-m-xylene

R.T.: 3.629 min
Delta R.T.: -0.006 min
Response: 23330344
Conc: 20.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.252 min
Delta R.T.: -0.012 min
Response: 40060691
Conc: 18.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.653 min
Delta R.T.: -0.015 min
Response: 17067234
Conc: 17.43 ng/ml



#31 AR-1260-1


R.T.: 7.222 min
Delta R.T.: -0.003 min
Response: 2703289
Conc: 20.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
B-30-SB01




#31 AR-1260-1

R.T.: 6.199 min
Delta R.T.: -0.011 min
Response: 977441
Conc: 16.08 ng/ml




#32 AR-1260-2

R.T.: 7.476 min
Delta R.T.: -0.009 min
Response: 4449330
Conc: 30.14 ng/ml



#32 AR-1260-2


R.T.: 6.388 min
Delta R.T.: -0.011 min
Response: 1315065
Conc: 18.19 ng/ml



#33 AR-1260-3


R.T.: 7.840 min
Delta R.T.: -0.008 min
Response: 2391727
Conc: 21.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
B-30-SB01




#33 AR-1260-3

R.T.: 6.541 min
Delta R.T.: -0.011 min
Response: 946063
Conc: 14.15 ng/ml




#34 AR-1260-4

R.T.: 8.062 min
Delta R.T.: -0.012 min
Response: 2989813
Conc: 23.31 ng/ml



#34 AR-1260-4


R.T.: 7.015 min
Delta R.T.: -0.012 min
Response: 735582
Conc: 13.37 ng/ml



#35 AR-1260-5

R.T.: 8.395 min
Delta R.T.: -0.007 min
Response: 3294187
Conc: 14.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
B-30-SB01



#35 AR-1260-5

R.T.: 7.259 min
Delta R.T.: -0.011 min
Response: 1451616
Conc: 12.33 ng/ml