

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032825\
 Data File : P0110100.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Mar 2025 16:31
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 18:04:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 19 02:52:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.691	3.688	441.1E6	252.7E6	48.462	48.191
2) SA Decachlor...	8.740	8.691	410.2E6	117.0E6	53.357	48.215
Target Compounds						
21) L5 AR-1248-1	4.782	4.769	104.7E6	57239126	474.919	470.960
22) L5 AR-1248-2	5.021	5.006	137.3E6	77819256	455.723	461.948
23) L5 AR-1248-3	5.236	5.048	183.7E6	85563490	492.180	472.232
24) L5 AR-1248-4	5.590	5.219	251.1E6	100.8E6	475.600	479.044
25) L5 AR-1248-5	5.632	5.610	179.6E6	95992475	481.492	461.786



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO032825\
Data File : PO110100.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Mar 2025 16:31
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 18:04:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: -0.001 min
Response: 441079565
Conc: 48.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.001 min
Response: 252718257
Conc: 48.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.740 min
Delta R.T.: -0.006 min
Response: 410246503
Conc: 53.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.691 min
Delta R.T.: -0.004 min
Response: 116975638
Conc: 48.22 ng/ml



#21 AR-1248-1


R.T.: 4.782 min
Delta R.T.: -0.003 min
Response: 104711229
Conc: 474.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.769 min
Delta R.T.: -0.003 min
Response: 57239126
Conc: 470.96 ng/ml




#22 AR-1248-2

R.T.: 5.021 min
Delta R.T.: -0.002 min
Response: 137258073
Conc: 455.72 ng/ml



#22 AR-1248-2


R.T.: 5.006 min
Delta R.T.: -0.002 min
Response: 77819256
Conc: 461.95 ng/ml



#23 AR-1248-3


R.T.: 5.236 min
Delta R.T.: -0.002 min
Response: 183675988
Conc: 492.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.048 min
Delta R.T.: -0.003 min
Response: 85563490
Conc: 472.23 ng/ml




#24 AR-1248-4

R.T.: 5.590 min
Delta R.T.: -0.002 min
Response: 251126122
Conc: 475.60 ng/ml



#24 AR-1248-4


R.T.: 5.219 min
Delta R.T.: -0.003 min
Response: 100818460
Conc: 479.04 ng/ml



#25 AR-1248-5

R.T.: 5.632 min
Delta R.T.: -0.003 min
Response: 179572292
Conc: 481.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.610 min
Delta R.T.: -0.003 min
Response: 95992475
Conc: 461.79 ng/ml