

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0032922\
 Data File : P0085714.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Mar 2022 17:23
 Operator : YP\AJ
 Sample : N2123-03MSD
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 SPLP-LEACHATE-AMENDED-COMP-CS-1MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/30/2022
 Supervised By :mohammad ahmed 03/30/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 03:20:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0032822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 29 06:17:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.464	3.564	3851673	952557	19.307m	19.814
2) SA Decachlor...	10.364	8.585	3539713	823596	26.993	23.161
Target Compounds						
3) L1 AR-1016-1	5.654	4.645	4517820	3989727	676.530m	2227.019 #
4) L1 AR-1016-2	5.677	4.671	5618393	1102046	555.823m	434.567
5) L1 AR-1016-3	5.741	4.847	3120846	660122	514.734m	474.623
6) L1 AR-1016-4	5.841	4.891	2793449	520467	572.196m	444.129
7) L1 AR-1016-5	6.140	5.103	2134127	604160	469.505m	446.013m
31) L7 AR-1260-1	7.284	6.142	4503869	1396685	562.295	559.342
32) L7 AR-1260-2	7.543	6.332	5315212	1577534	566.487	518.484
33) L7 AR-1260-3	7.908	6.484	3459182	1515820	490.280m	552.281
34) L7 AR-1260-4	8.138	6.959	4245102	1066641	539.141m	465.331
35) L7 AR-1260-5	8.468	7.204	8565533	2443813	557.559	483.373

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032922\
Data File : P0085714.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Mar 2022 17:23
Operator : YP\AJ
Sample : N2123-03MSD
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SPLP-LEACHATE-AMENDED-COMP-CS-1MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/30/2022
Supervised By :mohammad ahmed 03/30/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 03:20:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 29 06:17:49 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

