

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032922\
 Data File : P0085722.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Mar 2022 20:09
 Operator : YP\AJ
 Sample : N2155-08
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 22L076-H

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/30/2022
 Supervised By :mohammad ahmed 03/30/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 03:24:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 29 06:17:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.460	3.562	3705578	958858	18.575m	19.945
2) SA Decachlor...	10.351	8.577	2856711	686623	21.784	19.309

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032922\
Data File : P0085722.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Mar 2022 20:09
Operator : YP\AJ
Sample : N2155-08
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
22L076-H


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/30/2022
Supervised By :mohammad ahmed 03/30/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 03:24:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 29 06:17:49 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.460 min
Delta R.T.: 0.000 min
Response: 3705578
Conc: 18.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
22L076-H


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/30/2022
Supervised By :mohammad ahmed 03/30/2022




#1 Tetrachloro-m-xylene

R.T.: 3.562 min
Delta R.T.: 0.000 min
Response: 958858
Conc: 19.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.351 min
Delta R.T.: -0.005 min
Response: 2856711
Conc: 21.78 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.577 min
Delta R.T.: -0.002 min
Response: 686623
Conc: 19.31 ng/ml