

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032922\
 Data File : P0085753.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Mar 2022 5:40
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 06:40:04 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 29 06:17:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.467	3.567	10460892	2377966	52.436	49.463
2)	SA Decachlor...	10.373	8.591	6347010	1580087	48.400	44.434
Target Compounds							
16)	L4 AR-1242-1	5.659	4.656	2582234	651596	479.586	461.633
17)	L4 AR-1242-2	5.681	4.675	3652138	930291	449.849	477.179
18)	L4 AR-1242-3	5.745	4.850	2282325	504401	455.102	471.120
19)	L4 AR-1242-4	5.845	4.936	1876987	509067	474.549	469.246
20)	L4 AR-1242-5	6.593	5.460	1981350	673048	444.404	503.063



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032922\
Data File : P0085753.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Mar 2022 5:40
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 06:40:04 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 29 06:17:49 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.467 min
Delta R.T.: 0.006 min
Response: 10460892
Conc: 52.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.567 min
Delta R.T.: 0.006 min
Response: 2377966
Conc: 49.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.373 min
Delta R.T.: 0.017 min
Response: 6347010
Conc: 48.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.591 min
Delta R.T.: 0.012 min
Response: 1580087
Conc: 44.43 ng/ml



#16 AR-1242-1


R.T.: 5.659 min
Delta R.T.: 0.008 min
Response: 2582234
Conc: 479.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.656 min
Delta R.T.: 0.006 min
Response: 651596
Conc: 461.63 ng/ml




#17 AR-1242-2

R.T.: 5.681 min
Delta R.T.: 0.008 min
Response: 3652138
Conc: 449.85 ng/ml



#17 AR-1242-2


R.T.: 4.675 min
Delta R.T.: 0.006 min
Response: 930291
Conc: 477.18 ng/ml



#18 AR-1242-3


R.T.: 5.745 min
Delta R.T.: 0.009 min
Response: 2282325
Conc: 455.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.850 min
Delta R.T.: 0.006 min
Response: 504401
Conc: 471.12 ng/ml




#19 AR-1242-4

R.T.: 5.845 min
Delta R.T.: 0.008 min
Response: 1876987
Conc: 474.55 ng/ml



#19 AR-1242-4


R.T.: 4.936 min
Delta R.T.: 0.006 min
Response: 509067
Conc: 469.25 ng/ml



#20 AR-1242-5

R.T.: 6.593 min
Delta R.T.: 0.008 min
Response: 1981350
Conc: 444.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.460 min
Delta R.T.: 0.007 min
Response: 673048
Conc: 503.06 ng/ml