

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033021\
 Data File : P0076633.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Mar 2021 23:48
 Operator : DD\AJ
 Sample : M1855-01
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 EMBANKMENT-FILL-COMPOSITE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 31 02:10:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 24 05:38:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.912	3.925	2066702	1152712	23.598	21.014
2)	SA Decachlor...	10.884	9.180	1999651	1157602	30.719	24.566
Target Compounds							
3)	L1 AR-1016-1	6.261f	0.000	51591	0	19.090	N.D. #
4)	L1 AR-1016-2	6.261	0.000	51591	0	12.550	N.D. #
5)	L1 AR-1016-3	6.324	5.375	22992	22183	9.529	16.614 #
6)	L1 AR-1016-4	0.000	5.434	0	13253	N.D.	11.806 #
7)	L1 AR-1016-5	0.000	5.661	0	49257	N.D.	35.771 #
8)	L2 AR-1221-1	5.165	0.000	141116	0	160.386	N.D. #
9)	L2 AR-1221-2	5.266	4.274	28617	21985	42.257	51.021
10)	L2 AR-1221-3	5.363	0.000	45973	0	21.246	N.D. #
11)	L3 AR-1232-1	5.363	0.000	45973	0	25.989	N.D. #
12)	L3 AR-1232-2	5.936	0.000	63993	0	72.149	N.D. #
13)	L3 AR-1232-3	6.261	5.375	51591	22183	30.551	40.626 #
15)	L3 AR-1232-5	6.518	5.661	35252	49257	63.947	90.295 #
16)	L4 AR-1242-1	6.261f	0.000	51591	0	23.077	N.D. #
17)	L4 AR-1242-2	6.261	0.000	51591	0	15.184	N.D. #
18)	L4 AR-1242-3	6.324	5.375	22992	22183	11.377	20.487 #
20)	L4 AR-1242-5	7.232f	6.031	272336	247501	155.918	186.593
21)	L5 AR-1248-1	6.261f	0.000	51591	0	31.167	N.D. #
22)	L5 AR-1248-2	6.518	5.434	35252	13253	14.441	8.139 #
24)	L5 AR-1248-4	7.181	5.661	481869	49257	154.319	25.336 #
25)	L5 AR-1248-5	7.232f	6.083	272336	83105	87.460	44.850 #
26)	L6 AR-1254-1	7.151	6.031	330617	247501	99.539	83.561
27)	L6 AR-1254-2	7.378	6.184f	408069	94998	79.284	36.564 #
28)	L6 AR-1254-3	7.756	6.613	339157	63791	63.761	15.927 #
29)	L6 AR-1254-4	8.047	6.854	109452	53479	30.347	22.570 #
30)	L6 AR-1254-5	8.475	7.278	135737	81064	34.628	23.000 #
31)	L7 AR-1260-1	7.917	6.749	200807	64052	58.436	24.992 #
32)	L7 AR-1260-2	8.181	6.947	450790	53745	111.180	17.330 #
33)	L7 AR-1260-3	8.546	7.097	58870	66294	19.486	22.566
34)	L7 AR-1260-4	8.774	7.582	56916	29317	16.333	12.301
35)	L7 AR-1260-5	9.101	7.826	79353	69773	11.742	12.298
36)	L8 AR-1262-1	8.546	7.278	58870	81064	11.378	41.223 #
37)	L8 AR-1262-2	9.101	7.826	79353	69773	9.208	10.378
38)	L8 AR-1262-3	9.426f	0.000	59241	0	10.449	N.D. #
39)	L8 AR-1262-4	9.500	8.173	21252	39969	4.937	8.031 #
40)	L8 AR-1262-5	10.154	0.000	8863	0	3.050	N.D. #
41)	L9 AR-1268-1	9.426f	0.000	59241	0	5.710	N.D. #
42)	L9 AR-1268-2	9.500	8.173	21252	39969	2.284	5.635 #
43)	L9 AR-1268-3	9.728	0.000	13849	0	1.667	N.D. #
44)	L9 AR-1268-4	10.154	0.000	8863	0	2.727	N.D. #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033021\
 Data File : P0076633.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Mar 2021 23:48
 Operator : DD\AJ
 Sample : M1855-01
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 EMBANKMENT-FILL-COMPOSITE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 31 02:10:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 24 05:38:14 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
45) L9	AR-1268-5	10.558	0.000	57770	0	2.183	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033021\
Data File : P0076633.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Mar 2021 23:48
Operator : DD\AJ
Sample : M1855-01
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 02:10:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032421.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 24 05:38:14 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.912 min
Delta R.T.: 0.000 min
Response: 2066702
Conc: 23.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#1 Tetrachloro-m-xylene

R.T.: 3.925 min
Delta R.T.: -0.002 min
Response: 1152712
Conc: 21.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.884 min
Delta R.T.: -0.003 min
Response: 1999651
Conc: 30.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.180 min
Delta R.T.: -0.006 min
Response: 1157602
Conc: 24.57 ng/ml



#3 AR-1016-1


R.T.: 6.261 min
Delta R.T.: 0.029 min
Response: 51591
Conc: 19.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 5.175 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 6.261 min
Delta R.T.: 0.006 min
Response: 51591
Conc: 12.55 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 5.195 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 6.324 min
Delta R.T.: 0.002 min
Response: 22992
Conc: 9.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#5 AR-1016-3

R.T.: 5.375 min
Delta R.T.: -0.009 min
Response: 22183
Conc: 16.61 ng/ml




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T. : 6.429 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 5.434 min
Delta R.T.: -0.001 min
Response: 13253
Conc: 11.81 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 6.743 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#7 AR-1016-5

R.T.: 5.661 min
Delta R.T.: -0.002 min
Response: 49257
Conc: 35.77 ng/ml




#8 AR-1221-1

R.T.: 5.165 min
Delta R.T.: -0.002 min
Response: 141116
Conc: 160.39 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 4.187 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 5.266 min
Delta R.T.: 0.000 min
Response: 28617
Conc: 42.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#9 AR-1221-2

R.T.: 4.274 min
Delta R.T.: -0.011 min
Response: 21985
Conc: 51.02 ng/ml




#10 AR-1221-3

R.T.: 5.363 min
Delta R.T.: 0.011 min
Response: 45973
Conc: 21.25 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 4.373 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 5.363 min
Delta R.T.: 0.012 min
Response: 45973
Conc: 25.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.372 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.936 min
Delta R.T.: -0.003 min
Response: 63993
Conc: 72.15 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 5.197 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 6.261 min
Delta R.T.: 0.001 min
Response: 51591
Conc: 30.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#13 AR-1232-3

R.T.: 5.375 min
Delta R.T.: -0.011 min
Response: 22183
Conc: 40.63 ng/ml




#15 AR-1232-5

R.T.: 6.518 min
Delta R.T.: -0.012 min
Response: 35252
Conc: 63.95 ng/ml



#15 AR-1232-5


R.T.: 5.661 min
Delta R.T.: -0.004 min
Response: 49257
Conc: 90.30 ng/ml



#16 AR-1242-1


R.T.: 6.261 min
Delta R.T.: 0.026 min
Response: 51591
Conc: 23.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T. : 5.176 min
Response: 0
Conc: N.D.




#17 AR-1242-2

R.T.: 6.261 min
Delta R.T.: 0.003 min
Response: 51591
Conc: 15.18 ng/ml



#17 AR-1242-2


R.T.: 0.000 min
Exp R.T. : 5.196 min
Response: 0
Conc: N.D.



#18 AR-1242-3


R.T.: 6.324 min
Delta R.T.: 0.000 min
Response: 22992
Conc: 11.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#18 AR-1242-3

R.T.: 5.375 min
Delta R.T.: -0.011 min
Response: 22183
Conc: 20.49 ng/ml




#20 AR-1242-5

R.T.: 7.232 min
Delta R.T.: 0.018 min
Response: 272336
Conc: 155.92 ng/ml



#20 AR-1242-5


R.T.: 6.031 min
Delta R.T.: -0.011 min
Response: 247501
Conc: 186.59 ng/ml



#21 AR-1248-1


R.T.: 6.261 min
Delta R.T.: 0.026 min
Response: 51591
Conc: 31.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 5.177 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 6.518 min
Delta R.T.: -0.011 min
Response: 35252
Conc: 14.44 ng/ml



#22 AR-1248-2


R.T.: 5.434 min
Delta R.T.: -0.003 min
Response: 13253
Conc: 8.14 ng/ml



#24 AR-1248-4


R.T.: 7.181 min
Delta R.T.: 0.006 min
Response: 481869
Conc: 154.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#24 AR-1248-4

R.T.: 5.661 min
Delta R.T.: -0.003 min
Response: 49257
Conc: 25.34 ng/ml




#25 AR-1248-5

R.T.: 7.232 min
Delta R.T.: 0.017 min
Response: 272336
Conc: 87.46 ng/ml



#25 AR-1248-5


R.T.: 6.083 min
Delta R.T.: -0.003 min
Response: 83105
Conc: 44.85 ng/ml



#26 AR-1254-1


R.T.: 7.151 min
Delta R.T.: 0.006 min
Response: 330617
Conc: 99.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#26 AR-1254-1

R.T.: 6.031 min
Delta R.T.: -0.011 min
Response: 247501
Conc: 83.56 ng/ml




#27 AR-1254-2

R.T.: 7.378 min
Delta R.T.: 0.004 min
Response: 408069
Conc: 79.28 ng/ml



#27 AR-1254-2


R.T.: 6.184 min
Delta R.T.: -0.018 min
Response: 94998
Conc: 36.56 ng/ml



#28 AR-1254-3


R.T.: 7.756 min
Delta R.T.: 0.000 min
Response: 339157
Conc: 63.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#28 AR-1254-3

R.T.: 6.613 min
Delta R.T.: -0.006 min
Response: 63791
Conc: 15.93 ng/ml




#29 AR-1254-4

R.T.: 8.047 min
Delta R.T.: -0.005 min
Response: 109452
Conc: 30.35 ng/ml



#29 AR-1254-4


R.T.: 6.854 min
Delta R.T.: -0.005 min
Response: 53479
Conc: 22.57 ng/ml



#30 AR-1254-5


R.T.: 8.475 min
Delta R.T.: -0.004 min
Response: 135737
Conc: 34.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#30 AR-1254-5

R.T.: 7.278 min
Delta R.T.: -0.007 min
Response: 81064
Conc: 23.00 ng/ml




#31 AR-1260-1

R.T.: 7.917 min
Delta R.T.: 0.000 min
Response: 200807
Conc: 58.44 ng/ml



#31 AR-1260-1


R.T.: 6.749 min
Delta R.T.: -0.004 min
Response: 64052
Conc: 24.99 ng/ml



#32 AR-1260-2


R.T.: 8.181 min
Delta R.T.: 0.000 min
Response: 450790
Conc: 111.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#32 AR-1260-2

R.T.: 6.947 min
Delta R.T.: -0.003 min
Response: 53745
Conc: 17.33 ng/ml




#33 AR-1260-3

R.T.: 8.546 min
Delta R.T.: 0.000 min
Response: 58870
Conc: 19.49 ng/ml



#33 AR-1260-3


R.T.: 7.097 min
Delta R.T.: -0.006 min
Response: 66294
Conc: 22.57 ng/ml



#34 AR-1260-4


R.T.: 8.774 min
Delta R.T.: -0.002 min
Response: 56916
Conc: 16.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#34 AR-1260-4

R.T.: 7.582 min
Delta R.T.: -0.002 min
Response: 29317
Conc: 12.30 ng/ml




#35 AR-1260-5

R.T.: 9.101 min
Delta R.T.: 0.003 min
Response: 79353
Conc: 11.74 ng/ml



#35 AR-1260-5


R.T.: 7.826 min
Delta R.T.: -0.006 min
Response: 69773
Conc: 12.30 ng/ml



#36 AR-1262-1


R.T.: 8.546 min
Delta R.T.: -0.004 min
Response: 58870
Conc: 11.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#36 AR-1262-1

R.T.: 7.278 min
Delta R.T.: -0.003 min
Response: 81064
Conc: 41.22 ng/ml




#37 AR-1262-2

R.T.: 9.101 min
Delta R.T.: -0.002 min
Response: 79353
Conc: 9.21 ng/ml



#37 AR-1262-2


R.T.: 7.826 min
Delta R.T.: -0.006 min
Response: 69773
Conc: 10.38 ng/ml



#38 AR-1262-3


R.T.: 9.426 min
Delta R.T.: 0.016 min
Response: 59241
Conc: 10.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.117 min
Response: 0
Conc: N.D.




#39 AR-1262-4

R.T.: 9.500 min
Delta R.T.: 0.002 min
Response: 21252
Conc: 4.94 ng/ml



#39 AR-1262-4


R.T.: 8.173 min
Delta R.T.: -0.007 min
Response: 39969
Conc: 8.03 ng/ml



#40 AR-1262-5


R.T.: 10.154 min
Delta R.T.: -0.002 min
Response: 8863
Conc: 3.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 8.677 min
Response: 0
Conc: N.D.




#41 AR-1268-1

R.T.: 9.426 min
Delta R.T.: 0.017 min
Response: 59241
Conc: 5.71 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T. : 8.117 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 9.500 min
Delta R.T.: -0.004 min
Response: 21252
Conc: 2.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#42 AR-1268-2

R.T.: 8.173 min
Delta R.T.: -0.008 min
Response: 39969
Conc: 5.63 ng/ml




#43 AR-1268-3

R.T.: 9.728 min
Delta R.T.: 0.004 min
Response: 13849
Conc: 1.67 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 8.387 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 10.154 min
Delta R.T.: -0.002 min
Response: 8863
Conc: 2.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
EMBANKMENT-FILL-COMPOSITE




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T. : 8.676 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 10.558 min
Delta R.T.: -0.004 min
Response: 57770
Conc: 2.18 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 8.950 min
Response: 0
Conc: N.D.