

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033022\
 Data File : P0085755.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Mar 2022 12:17
 Operator : YP\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 16:10:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 30 16:02:31 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds							
21)	L5	AR-1248-1	5.626f	4.654	15149	7048	4.013 7.385 #
22)	L5	AR-1248-2	5.933	4.897	7660	4134	1.440 2.935 #
23)	L5	AR-1248-3	0.000	4.897f	0	4134	N.D. 2.896 #
24)	L5	AR-1248-4	6.559	5.139	103006	10925	20.313 6.442 #
25)	L5	AR-1248-5	6.597	5.507	46544	11025	7.356 6.760
26)	L6	AR-1254-1	6.559	5.507f	103006	11025	14.817 4.262 #
27)	L6	AR-1254-2	6.745	0.000	24132	0	2.253 N.D. #
28)	L6	AR-1254-3	7.117	0.000	3757	0	0.336 N.D. #
29)	L6	AR-1254-4	7.415	0.000	35258	0	4.425 N.D. #
30)	L6	AR-1254-5	7.838	0.000	26354	0	3.262 N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033022\
Data File : P0085755.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Mar 2022 12:17
Operator : YP\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 16:10:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 30 16:02:31 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 4.475 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE

#1 Tetrachloro-m-xylene


R.T.: 0.000 min
Exp R.T. : 3.573 min
Response: 0
Conc: N.D.

#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.380 min
Response: 0
Conc: N.D.

#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 8.596 min
Response: 0
Conc: N.D.



#21 AR-1248-1


R.T.: 5.626 min
Delta R.T.: -0.040 min
Response: 15149
Conc: 4.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#21 AR-1248-1

R.T.: 4.654 min
Delta R.T.: -0.009 min
Response: 7048
Conc: 7.39 ng/ml




#22 AR-1248-2

R.T.: 5.933 min
Delta R.T.: -0.011 min
Response: 7660
Conc: 1.44 ng/ml



#22 AR-1248-2


R.T.: 4.897 min
Delta R.T.: -0.004 min
Response: 4134
Conc: 2.93 ng/ml



#23 AR-1248-3


R.T.: 0.000 min
Exp R.T.: 6.152 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
HEXANE




#23 AR-1248-3

R.T.: 4.897 min
Delta R.T.: -0.045 min
Response: 4134
Conc: 2.90 ng/ml




#24 AR-1248-4

R.T.: 6.559 min
Delta R.T.: 0.025 min
Response: 103006
Conc: 20.31 ng/ml



#24 AR-1248-4


R.T.: 5.139 min
Delta R.T.: 0.025 min
Response: 10925
Conc: 6.44 ng/ml



#25 AR-1248-5


R.T.: 6.597 min
Delta R.T.: -0.003 min
Response: 46544
Conc: 7.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#25 AR-1248-5

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 11025
Conc: 6.76 ng/ml




#26 AR-1254-1

R.T.: 6.559 min
Delta R.T.: 0.019 min
Response: 103006
Conc: 14.82 ng/ml



#26 AR-1254-1


R.T.: 5.507 min
Delta R.T.: 0.036 min
Response: 11025
Conc: 4.26 ng/ml



#27 AR-1254-2


R.T.: 6.745 min
Delta R.T.: -0.017 min
Response: 24132
Conc: 2.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.621 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.117 min
Delta R.T.: -0.016 min
Response: 3757
Conc: 0.34 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.024 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.415 min
Delta R.T.: -0.007 min
Response: 35258
Conc: 4.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEXANE




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.255 min
Response: 0
Conc: N.D.



#30 AR-1254-5

R.T.: 7.838 min
Delta R.T.: -0.008 min
Response: 26354
Conc: 3.26 ng/ml



#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 6.674 min
Response: 0
Conc: N.D.