

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033022\
 Data File : P0085760.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Mar 2022 13:37
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 13:51:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 30 13:43:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.479	3.579	5136646	1209594	24.444	25.107
2) SA Decachlor...	10.382	8.596	3405090	879051	25.214	27.091
Target Compounds						
21) L5 AR-1248-1	5.669	4.666	1020522	249755	259.924	256.855
22) L5 AR-1248-2	5.947	4.904	1476711	382532	264.925	280.287
23) L5 AR-1248-3	6.154	4.945	1632323	397150	262.446	277.892
24) L5 AR-1248-4	6.536	5.116	1370087	461840	259.832	274.679
25) L5 AR-1248-5	6.603	5.510	1709762	450234	258.010	270.715



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033022\
Data File : P0085760.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Mar 2022 13:37
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 13:51:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 30 13:43:07 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 5136646
Conc: 24.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.579 min
Delta R.T.: 0.001 min
Response: 1209594
Conc: 25.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.382 min
Delta R.T.: -0.002 min
Response: 3405090
Conc: 25.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.596 min
Delta R.T.: -0.001 min
Response: 879051
Conc: 27.09 ng/ml



#21 AR-1248-1


R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 1020522
Conc: 259.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 4.666 min
Delta R.T.: 0.000 min
Response: 249755
Conc: 256.85 ng/ml




#22 AR-1248-2

R.T.: 5.947 min
Delta R.T.: 0.000 min
Response: 1476711
Conc: 264.92 ng/ml



#22 AR-1248-2


R.T.: 4.904 min
Delta R.T.: 0.000 min
Response: 382532
Conc: 280.29 ng/ml



#23 AR-1248-3


R.T.: 6.154 min
Delta R.T.: 0.000 min
Response: 1632323
Conc: 262.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 4.945 min
Delta R.T.: 0.000 min
Response: 397150
Conc: 277.89 ng/ml




#24 AR-1248-4

R.T.: 6.536 min
Delta R.T.: 0.000 min
Response: 1370087
Conc: 259.83 ng/ml



#24 AR-1248-4


R.T.: 5.116 min
Delta R.T.: 0.000 min
Response: 461840
Conc: 274.68 ng/ml



#25 AR-1248-5

R.T.: 6.603 min
Delta R.T.: 0.000 min
Response: 1709762
Conc: 258.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 5.510 min
Delta R.T.: 0.000 min
Response: 450234
Conc: 270.72 ng/ml