

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033022\
 Data File : P0085785.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Mar 2022 21:55
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 31 04:22:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 30 16:55:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.465	3.564	9252821	2102373	47.482	44.452
2) SA Decachlor...	10.373	8.589	6282976	1530213	48.964	43.583
Target Compounds						
3) L1 AR-1016-1	5.655	4.653	3160504	782451	473.276	436.755
4) L1 AR-1016-2	5.678	4.672	4525844	1108922	447.738	437.279
5) L1 AR-1016-3	5.741	4.847	2770667	599140	456.977	430.777
6) L1 AR-1016-4	5.841	4.891	2274030	510890	465.801	435.956
7) L1 AR-1016-5	6.142	5.104	2247254	641819	494.393	473.814
31) L7 AR-1260-1	7.286	6.142	3687749	1101993	460.404	441.325
32) L7 AR-1260-2	7.546	6.333	4298582	1317232	458.136	432.931
33) L7 AR-1260-3	7.912	6.486	3272012	1226543	463.752	446.885
34) L7 AR-1260-4	8.143	6.961	3734690	1023182	474.317	446.371
35) L7 AR-1260-5	8.473	7.206	7048249	2304071	458.794	455.733



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033022\
Data File : P0085785.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Mar 2022 21:55
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 04:22:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 30 16:55:12 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.465 min
Delta R.T.: 0.004 min
Response: 9252821
Conc: 47.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.564 min
Delta R.T.: 0.004 min
Response: 2102373
Conc: 44.45 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.373 min
Delta R.T.: 0.015 min
Response: 6282976
Conc: 48.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.589 min
Delta R.T.: 0.010 min
Response: 1530213
Conc: 43.58 ng/ml



#3 AR-1016-1


R.T.: 5.655 min
Delta R.T.: 0.006 min
Response: 3160504
Conc: 473.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.653 min
Delta R.T.: 0.003 min
Response: 782451
Conc: 436.76 ng/ml




#4 AR-1016-2

R.T.: 5.678 min
Delta R.T.: 0.005 min
Response: 4525844
Conc: 447.74 ng/ml



#4 AR-1016-2


R.T.: 4.672 min
Delta R.T.: 0.003 min
Response: 1108922
Conc: 437.28 ng/ml



#5 AR-1016-3


R.T.: 5.741 min
Delta R.T.: 0.005 min
Response: 2770667
Conc: 456.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.847 min
Delta R.T.: 0.004 min
Response: 599140
Conc: 430.78 ng/ml




#6 AR-1016-4

R.T.: 5.841 min
Delta R.T.: 0.006 min
Response: 2274030
Conc: 465.80 ng/ml



#6 AR-1016-4


R.T.: 4.891 min
Delta R.T.: 0.004 min
Response: 510890
Conc: 435.96 ng/ml



#7 AR-1016-5


R.T.: 6.142 min
Delta R.T.: 0.006 min
Response: 2247254
Conc: 494.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.104 min
Delta R.T.: 0.000 min
Response: 641819
Conc: 473.81 ng/ml




#31 AR-1260-1

R.T.: 7.286 min
Delta R.T.: 0.007 min
Response: 3687749
Conc: 460.40 ng/ml



#31 AR-1260-1


R.T.: 6.142 min
Delta R.T.: 0.006 min
Response: 1101993
Conc: 441.32 ng/ml



#32 AR-1260-2


R.T.: 7.546 min
Delta R.T.: 0.008 min
Response: 4298582
Conc: 458.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.333 min
Delta R.T.: 0.007 min
Response: 1317232
Conc: 432.93 ng/ml




#33 AR-1260-3

R.T.: 7.912 min
Delta R.T.: 0.009 min
Response: 3272012
Conc: 463.75 ng/ml



#33 AR-1260-3


R.T.: 6.486 min
Delta R.T.: 0.007 min
Response: 1226543
Conc: 446.88 ng/ml



#34 AR-1260-4


R.T.: 8.143 min
Delta R.T.: 0.010 min
Response: 3734690
Conc: 474.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.961 min
Delta R.T.: 0.007 min
Response: 1023182
Conc: 446.37 ng/ml



#35 AR-1260-5

R.T.: 8.473 min
Delta R.T.: 0.009 min
Response: 7048249
Conc: 458.79 ng/ml



#35 AR-1260-5

R.T.: 7.206 min
Delta R.T.: 0.007 min
Response: 2304071
Conc: 455.73 ng/ml