

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033023\
 Data File : P0093660.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2023 11:36
 Operator : YP/AJ
 Sample : PB151792BL
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB151792BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 21:12:17 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 14 05:32:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.422	3.625	66389985	22135464	21.201	19.626
2) SA Decachlor...	10.250	8.650	52780789	22414405	24.387	22.892

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0033023\
Data File : P0093660.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Mar 2023 11:36
Operator : YP/AJ
Sample : PB151792BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB151792BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 21:12:17 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 14 05:32:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.422 min
Delta R.T.: -0.008 min
Response: 66389985
Conc: 21.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB151792BL




#1 Tetrachloro-m-xylene

R.T.: 3.625 min
Delta R.T.: -0.009 min
Response: 22135464
Conc: 19.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.250 min
Delta R.T.: -0.013 min
Response: 52780789
Conc: 24.39 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.650 min
Delta R.T.: -0.018 min
Response: 22414405
Conc: 22.89 ng/ml