

CHEMTECH 284 Sheffield Street, Mountainside NJ 07092 (908) 789 - 8900
Daily Maintenance Log :ECD_D
Date :Monday, April 3, 2023

Sequence ID :	PD010621	
Initials:	Ankita	
Date :	1/6/2021	
Routine Maintenance		
	Replace Septum	YES
	Clip Column	NO
	Inches	0
	Replace Inlet Liner	YES
	Replace O - Ring	YES
	Replace Ferrule	NO
	replaceColumn	NO
	Replace Trap VOC	NO
	Replaced Water Management VOC	NO
	Helium Lot #	
	Clean Source	NO
	Filamets Replaced	NO
	Replaced Rough Pump Oil	NO
Comment : baked oven at 300 degrees for 40 mins		
Sequence ID :	pd012521	
Initials:	Ankita	
Date :	1/25/2021	
Routine Maintenance		
	Replace Septum	YES
	Clip Column	NO
	Inches	0
	Replace Inlet Liner	YES
	Replace O - Ring	YES
	Replace Ferrule	NO
	replaceColumn	NO
	Replace Trap VOC	NO
	Replaced Water Management VOC	NO
	Helium Lot #	
	Clean Source	NO
	Filamets Replaced	NO
	Replaced Rough Pump Oil	NO
Comment : baked oven at 300 degrees for 40 mins		