

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0040121\
 Data File : P0076750.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Apr 2021 00:56
 Operator : DD\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 ICVPO040121AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 05:26:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0040121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 02 05:20:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	93.014	92.556 E3	0.5	101	0.00
2 SA Decachlorobiphenyl	92.620	160.287 E3	-73.1#	174#	0.00
41 L9 AR-1268-1	12.177	12.780 E3	-5.0	102	0.00
42 L9 AR-1268-2	11.306	11.804 E3	-4.4	102	0.00
43 L9 AR-1268-3	9.769	10.154 E3	-3.9	101	0.00
44 L9 AR-1268-4	3.909	4.089 E3	-4.6	102	0.00
45 L9 AR-1268-5	32.853	34.153 E3	-4.0	102	0.00

Signal #2

1 SA Tetrachloro-m-xylene	50.183	48.673 E3	3.0	98	0.00
2 SA Decachlorobiphenyl	48.526	74.846 E3	-54.2#	152#	0.00
41 L9 AR-1268-1	7.351	7.432 E3	-1.1	101	0.00
42 L9 AR-1268-2	6.681	6.937 E3	-3.8	104	0.00
43 L9 AR-1268-3	5.792	5.876 E3	-1.5	101	0.00
44 L9 AR-1268-4	2.209	2.239 E3	-1.4	101	0.00
45 L9 AR-1268-5	16.501	16.580 E3	-0.5	100	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1 AR-1016-1	2.907	0.000 E3	100.0#	0#	-6.22#
4 L1 AR-1016-2	4.713	0.000 E3	100.0#	0#	-6.25#
5 L1 AR-1016-3	2.853	0.000 E3	100.0#	0#	-6.31#
6 L1 AR-1016-4	2.274	0.000 E3	100.0#	0#	-6.42#
7 L1 AR-1016-5	2.169	0.000 E3	100.0#	0#	-6.73#
8 L2 AR-1221-1	904.764	0.000	100.0#	0#	-5.15#
9 L2 AR-1221-2	671.936	0.000	100.0#	0#	-5.25#
10 L2 AR-1221-3	2.066	0.000 E3	100.0#	0#	-5.34#
11 L3 AR-1232-1	1.648	0.000 E3	100.0#	0#	-5.34#
12 L3 AR-1232-2	903.466	0.000	100.0#	0#	-5.93#
13 L3 AR-1232-3	1.844	0.000 E3	100.0#	0#	-6.25#
14 L3 AR-1232-4	904.971	0.000	100.0#	0#	-6.42#
15 L3 AR-1232-5	571.998	0.000	100.0#	0#	-6.52#
16 L4 AR-1242-1	2.111	0.000 E3	100.0#	0#	-6.22#
17 L4 AR-1242-2	3.523	0.000 E3	100.0#	0#	-6.25#
18 L4 AR-1242-3	2.221	0.000 E3	100.0#	0#	-6.31#
19 L4 AR-1242-4	1.663	0.000 E3	100.0#	0#	-6.42#
20 L4 AR-1242-5	1.958	0.000 E3	100.0#	0#	-7.20#
21 L5 AR-1248-1	1.519	0.000 E3	100.0#	0#	-6.22#
22 L5 AR-1248-2	2.315	0.000 E3	100.0#	0#	-6.52#
23 L5 AR-1248-3	2.812	0.000 E3	100.0#	0#	-6.73#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0040121\
 Data File : P0076750.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Apr 2021 00:56
 Operator : DD\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO040121AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 05:26:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0040121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 02 05:20:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	3.371	0.000 E3	100.0#	0#	-7.16#
25 L5 AR-1248-5	3.447	0.000 E3	100.0#	0#	-7.20#
26 L6 AR-1254-1	3.179	0.000 E3	100.0#	0#	-7.13#
27 L6 AR-1254-2	5.124	0.000 E3	100.0#	0#	-7.36#
28 L6 AR-1254-3	5.511	0.000 E3	100.0#	0#	-7.74#
29 L6 AR-1254-4	3.937	0.000 E3	100.0#	0#	-8.04#
30 L6 AR-1254-5	4.398	0.000 E3	100.0#	0#	-8.47#
31 L7 AR-1260-1	4.005	0.000 E3	100.0#	0#	-7.91#
32 L7 AR-1260-2	4.953	0.000 E3	100.0#	0#	-8.18#
33 L7 AR-1260-3	3.883	0.000 E3	100.0#	0#	-8.54#
34 L7 AR-1260-4	4.307	0.000 E3	100.0#	0#	-8.77#
35 L7 AR-1260-5	8.522	0.000 E3	100.0#	0#	-9.09#
36 L8 AR-1262-1	5.542	0.000 E3	100.0#	0#	-8.54#
37 L8 AR-1262-2	9.784	0.000 E3	100.0#	0#	-9.09#
38 L8 AR-1262-3	6.452	0.000 E3	100.0#	0#	-9.40#
39 L8 AR-1262-4	5.014	0.000 E3	100.0#	0#	-9.49#
40 L8 AR-1262-5	3.458	0.000 E3	100.0#	0#	-10.14#

Signal #2

3 L1 AR-1016-1	1.401	0.000 E3	100.0#	0#	-5.14#
4 L1 AR-1016-2	2.827	0.000 E3	100.0#	0#	-5.16#
5 L1 AR-1016-3	1.311	0.000 E3	100.0#	0#	-5.35#
6 L1 AR-1016-4	1.172	0.000 E3	100.0#	0#	-5.40#
7 L1 AR-1016-5	1.347	0.000 E3	100.0#	0#	-5.63#
8 L2 AR-1221-1	492.419	0.000	100.0#	0#	-4.16#
9 L2 AR-1221-2	352.652	0.000	100.0#	0#	-4.25#
10 L2 AR-1221-3	1.134	0.000 E3	100.0#	0#	-4.34#
11 L3 AR-1232-1	924.415	0.000	100.0#	0#	-4.34#
12 L3 AR-1232-2	1.140	0.000 E3	100.0#	0#	-5.17#
13 L3 AR-1232-3	514.813	0.000	100.0#	0#	-5.36#
14 L3 AR-1232-4	470.123	0.000	100.0#	0#	-5.45#
15 L3 AR-1232-5	496.643	0.000	100.0#	0#	-5.63#
16 L4 AR-1242-1	0.990	0.000 E3	100.0#	0#	-5.15#
17 L4 AR-1242-2	2.129	0.000 E3	100.0#	0#	-5.17#
18 L4 AR-1242-3	931.892	0.000	100.0#	0#	-5.36#
19 L4 AR-1242-4	967.172	0.000	100.0#	0#	-5.45#
20 L4 AR-1242-5	1.206	0.000 E3	100.0#	0#	-6.01#
21 L5 AR-1248-1	753.119	0.000	100.0#	0#	-5.15#
22 L5 AR-1248-2	1.452	0.000 E3	100.0#	0#	-5.40#
23 L5 AR-1248-3	1.456	0.000 E3	100.0#	0#	-5.45#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040121\
 Data File : P0076750.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Apr 2021 00:56
 Operator : DD\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO040121AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 05:26:35 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 02 05:20:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	1.707	0.000 E3	100.0#	0#	-5.63#
25 L5 AR-1248-5	2.010	0.000 E3	100.0#	0#	-6.05#
26 L6 AR-1254-1	2.621	0.000 E3	100.0#	0#	-6.01#
27 L6 AR-1254-2	2.290	0.000 E3	100.0#	0#	-6.17#
28 L6 AR-1254-3	3.534	0.000 E3	100.0#	0#	-6.58#
29 L6 AR-1254-4	2.099	0.000 E3	100.0#	0#	-6.82#
30 L6 AR-1254-5	3.260	0.000 E3	100.0#	0#	-7.25#
31 L7 AR-1260-1	2.613	0.000 E3	100.0#	0#	-6.72#
32 L7 AR-1260-2	3.158	0.000 E3	100.0#	0#	-6.92#
33 L7 AR-1260-3	2.922	0.000 E3	100.0#	0#	-7.07#
34 L7 AR-1260-4	2.502	0.000 E3	100.0#	0#	-7.55#
35 L7 AR-1260-5	5.899	0.000 E3	100.0#	0#	-7.80#
36 L8 AR-1262-1	1.770	0.000 E3	100.0#	0#	-7.25#
37 L8 AR-1262-2	6.056	0.000 E3	100.0#	0#	-7.80#
38 L8 AR-1262-3	2.515	0.000 E3	100.0#	0#	-8.08#
39 L8 AR-1262-4	4.533	0.000 E3	100.0#	0#	-8.14#
40 L8 AR-1262-5	1.981	0.000 E3	100.0#	0#	-8.64#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0