

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040121\
 Data File : P0076746.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Apr 2021 23:47
 Operator : DD\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO040121AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 05:23:25 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 02 05:20:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.899	3.897	4382431	2344313	47.116	46.716
2)	SA Decachlor...	10.872	9.144	4365088	2108603	47.129	43.453
Target Compounds							
16)	L4 AR-1242-1	6.224	5.146	1066166	477813	505.098	482.660
17)	L4 AR-1242-2	6.248	5.166	1750775	996755	497.020	468.187
18)	L4 AR-1242-3	6.314	5.355	1117170	439626	503.060	471.756
19)	L4 AR-1242-4	6.422	5.448	856573	482044	514.956	498.405
20)	L4 AR-1242-5	7.204	6.010	1008382	624837	515.134	518.066



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040121\
Data File : P0076746.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Apr 2021 23:47
Operator : DD\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO040121AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 05:23:25 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040121.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 02 05:20:17 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 4382431
Conc: 47.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO040121AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.897 min
Delta R.T.: 0.000 min
Response: 2344313
Conc: 46.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.872 min
Delta R.T.: -0.004 min
Response: 4365088
Conc: 47.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.144 min
Delta R.T.: -0.003 min
Response: 2108603
Conc: 43.45 ng/ml



#16 AR-1242-1


R.T.: 6.224 min
Delta R.T.: 0.000 min
Response: 1066166
Conc: 505.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO040121AR1242




#16 AR-1242-1

R.T.: 5.146 min
Delta R.T.: 0.000 min
Response: 477813
Conc: 482.66 ng/ml




#17 AR-1242-2

R.T.: 6.248 min
Delta R.T.: 0.000 min
Response: 1750775
Conc: 497.02 ng/ml



#17 AR-1242-2


R.T.: 5.166 min
Delta R.T.: 0.000 min
Response: 996755
Conc: 468.19 ng/ml



#18 AR-1242-3


R.T.: 6.314 min
Delta R.T.: 0.000 min
Response: 1117170
Conc: 503.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO040121AR1242




#18 AR-1242-3

R.T.: 5.355 min
Delta R.T.: 0.000 min
Response: 439626
Conc: 471.76 ng/ml




#19 AR-1242-4

R.T.: 6.422 min
Delta R.T.: 0.000 min
Response: 856573
Conc: 514.96 ng/ml



#19 AR-1242-4


R.T.: 5.448 min
Delta R.T.: 0.000 min
Response: 482044
Conc: 498.41 ng/ml



#20 AR-1242-5

R.T.: 7.204 min
Delta R.T.: 0.000 min
Response: 1008382
Conc: 515.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO040121AR1242



#20 AR-1242-5

R.T.: 6.010 min
Delta R.T.: 0.000 min
Response: 624837
Conc: 518.07 ng/ml