

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040121\
 Data File : P0076751.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Apr 2021 9:18
 Operator : DD\AJ
 Sample : P0040121ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO040121

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 09:46:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 02 05:20:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.899	3.896	4559260	2390153	49.017	47.629
2) SA Decachlor...	10.870	9.142	4551227	2132656	49.139	43.949
Target Compounds						
3) L1 AR-1016-1	6.223	5.144	1444370	637106	496.867	454.797
4) L1 AR-1016-2	6.247	5.164	2324372	1327600	493.234	469.567
5) L1 AR-1016-3	6.312	5.353	1462289	608737	512.622	464.279
6) L1 AR-1016-4	6.420	5.402	1141228	582633	501.806	497.147
7) L1 AR-1016-5	6.733	5.629	1114840	654759	514.088	486.090
31) L7 AR-1260-1	7.908	6.716	2093231	1254785	522.685	480.216
32) L7 AR-1260-2	8.174	6.913	2540113	1506260	512.813	477.015
33) L7 AR-1260-3	8.538	7.066	1974470	1409471	508.485	482.424
34) L7 AR-1260-4	8.769	7.545	2237587	1196596	519.543	478.322
35) L7 AR-1260-5	9.092	7.794	4507264	2787328	528.905	472.539



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040121\
Data File : P0076751.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Apr 2021 9:18
Operator : DD\AJ
Sample : P0040121ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO040121

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 09:46:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040121.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 02 05:20:17 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.899 min
Delta R.T.: -0.001 min
Response: 4559260
Conc: 49.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO040121




#1 Tetrachloro-m-xylene

R.T.: 3.896 min
Delta R.T.: -0.001 min
Response: 2390153
Conc: 47.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.870 min
Delta R.T.: -0.005 min
Response: 4551227
Conc: 49.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.142 min
Delta R.T.: -0.006 min
Response: 2132656
Conc: 43.95 ng/ml



#3 AR-1016-1


R.T.: 6.223 min
Delta R.T.: 0.000 min
Response: 1444370
Conc: 496.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO040121




#3 AR-1016-1

R.T.: 5.144 min
Delta R.T.: 0.000 min
Response: 637106
Conc: 454.80 ng/ml




#4 AR-1016-2

R.T.: 6.247 min
Delta R.T.: 0.000 min
Response: 2324372
Conc: 493.23 ng/ml



#4 AR-1016-2


R.T.: 5.164 min
Delta R.T.: 0.000 min
Response: 1327600
Conc: 469.57 ng/ml



#5 AR-1016-3


R.T.: 6.312 min
Delta R.T.: -0.001 min
Response: 1462289
Conc: 512.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO040121




#5 AR-1016-3

R.T.: 5.353 min
Delta R.T.: -0.002 min
Response: 608737
Conc: 464.28 ng/ml




#6 AR-1016-4

R.T.: 6.420 min
Delta R.T.: 0.000 min
Response: 1141228
Conc: 501.81 ng/ml



#6 AR-1016-4


R.T.: 5.402 min
Delta R.T.: -0.002 min
Response: 582633
Conc: 497.15 ng/ml



#7 AR-1016-5


R.T.: 6.733 min
Delta R.T.: -0.001 min
Response: 1114840
Conc: 514.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO040121




#7 AR-1016-5

R.T.: 5.629 min
Delta R.T.: -0.002 min
Response: 654759
Conc: 486.09 ng/ml




#31 AR-1260-1

R.T.: 7.908 min
Delta R.T.: -0.001 min
Response: 2093231
Conc: 522.68 ng/ml



#31 AR-1260-1


R.T.: 6.716 min
Delta R.T.: -0.003 min
Response: 1254785
Conc: 480.22 ng/ml



#32 AR-1260-2


R.T.: 8.174 min
Delta R.T.: -0.002 min
Response: 2540113
Conc: 512.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO040121




#32 AR-1260-2

R.T.: 6.913 min
Delta R.T.: -0.004 min
Response: 1506260
Conc: 477.01 ng/ml




#33 AR-1260-3

R.T.: 8.538 min
Delta R.T.: -0.003 min
Response: 1974470
Conc: 508.48 ng/ml



#33 AR-1260-3


R.T.: 7.066 min
Delta R.T.: -0.004 min
Response: 1409471
Conc: 482.42 ng/ml



#34 AR-1260-4


R.T.: 8.769 min
Delta R.T.: -0.003 min
Response: 2237587
Conc: 519.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO040121




#34 AR-1260-4

R.T.: 7.545 min
Delta R.T.: -0.004 min
Response: 1196596
Conc: 478.32 ng/ml



#35 AR-1260-5

R.T.: 9.092 min
Delta R.T.: -0.002 min
Response: 4507264
Conc: 528.90 ng/ml



#35 AR-1260-5

R.T.: 7.794 min
Delta R.T.: -0.004 min
Response: 2787328
Conc: 472.54 ng/ml