

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040122\
 Data File : P0085831.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Apr 2022 20:38
 Operator : YP\AJ
 Sample : N2220-09MSD
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 A02145MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/04/2022
 Supervised By :mohammad ahmed 04/04/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 09:40:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 30 16:55:12 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.483	3.581	4492897	1013915	23.056	21.438
2) SA Decachlor...	10.388	8.601	2908012	749683	22.662	21.352
Target Compounds						
3) L1 AR-1016-1	5.672	4.670	3663876	904001	548.654	504.603
4) L1 AR-1016-2	5.695	4.688	5281649	1307002	522.509	515.387
5) L1 AR-1016-3	5.758	4.864	3220428	697873	531.158	501.765
6) L1 AR-1016-4	5.858	4.907	2598215	557940	532.205m	476.106
7) L1 AR-1016-5	6.158	5.120	2515630	725247	553.435	535.404
31) L7 AR-1260-1	7.301	6.157	4264341	1346397	532.390	539.203
32) L7 AR-1260-2	7.560	6.347	5152767	1603760	549.174	527.104
33) L7 AR-1260-3	7.925	6.499	3281164	1516921	465.049	552.682
34) L7 AR-1260-4	8.155	6.974	4039652	1073638	513.048	468.383
35) L7 AR-1260-5	8.486	7.218	7249373	2427634	471.886m	480.173

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040122\
Data File : P0085831.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Apr 2022 20:38
Operator : YP\AJ
Sample : N2220-09MSD
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
A02145MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/04/2022
Supervised By :mohammad ahmed 04/04/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 09:40:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 30 16:55:12 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

