

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040125\
 Data File : P0110140.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2025 09:17
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 10:44:44 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 19 02:52:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.691	3.688	424.8E6	239.6E6	46.676	45.690
2) SA Decachlor...	8.741	8.692	431.3E6	114.0E6	56.096	46.970
Target Compounds						
16) L4 AR-1242-1	4.783	4.770	136.1E6	71275091	508.834	487.466
17) L4 AR-1242-2	4.803	4.789	188.9E6	101.7E6	512.932	489.013
18) L4 AR-1242-3	4.859	4.964	132.2E6	55410119	512.396	496.449
19) L4 AR-1242-4	4.980	5.049	103.3E6	56150956	516.340	484.158
20) L4 AR-1242-5	5.633	5.570	117.3E6	70165974	555.899	511.257



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040125\
Data File : PO110140.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Apr 2025 09:17
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 01 10:44:44 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031925.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 19 02:52:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.691 min
Delta R.T.: 0.000 min
Response: 424822376
Conc: 46.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.688 min
Delta R.T.: -0.002 min
Response: 239604152
Conc: 45.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.741 min
Delta R.T.: -0.004 min
Response: 431303319
Conc: 56.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.692 min
Delta R.T.: -0.003 min
Response: 113955084
Conc: 46.97 ng/ml



#16 AR-1242-1


R.T.: 4.783 min
Delta R.T.: -0.002 min
Response: 136065852
Conc: 508.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.770 min
Delta R.T.: -0.002 min
Response: 71275091
Conc: 487.47 ng/ml




#17 AR-1242-2

R.T.: 4.803 min
Delta R.T.: -0.002 min
Response: 188863084
Conc: 512.93 ng/ml



#17 AR-1242-2


R.T.: 4.789 min
Delta R.T.: -0.002 min
Response: 101731990
Conc: 489.01 ng/ml



#18 AR-1242-3


R.T.: 4.859 min
Delta R.T.: -0.002 min
Response: 132173022
Conc: 512.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.964 min
Delta R.T.: -0.003 min
Response: 55410119
Conc: 496.45 ng/ml




#19 AR-1242-4

R.T.: 4.980 min
Delta R.T.: -0.002 min
Response: 103316562
Conc: 516.34 ng/ml



#19 AR-1242-4


R.T.: 5.049 min
Delta R.T.: -0.003 min
Response: 56150956
Conc: 484.16 ng/ml



#20 AR-1242-5

R.T.: 5.633 min
Delta R.T.: -0.002 min
Response: 117325779
Conc: 555.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.570 min
Delta R.T.: -0.003 min
Response: 70165974
Conc: 511.26 ng/ml