

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040221\
 Data File : P0076765.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Apr 2021 13:55
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Ankita
 4/5/2021 3:32:45 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 16:05:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 02 05:20:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.899	3.896	4794715	2489739	51.549	49.614
2) SA Decachlor...	10.871	9.142	4738859	2205855	51.165	45.457
Target Compounds						
26) L6 AR-1254-1	7.130	6.005	1769834	1385408	556.809m	528.554m
27) L6 AR-1254-2	7.360	6.165	2923531	1226111	570.593	535.417m
28) L6 AR-1254-3	7.742	6.582	3289650	1980726	596.972	560.432m
29) L6 AR-1254-4	8.039	6.823	2416853	1229140	613.839	585.671
30) L6 AR-1254-5	8.465	7.246	2512172	1778526	571.192	545.558

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040221\
Data File : P0076765.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Apr 2021 13:55
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
4/5/2021 3:32:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 16:05:53 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040121.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 02 05:20:17 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.899 min
Delta R.T.: -0.001 min
Response: 4794715
Conc: 51.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
4/5/2021 3:32:45 PM




#1 Tetrachloro-m-xylene

R.T.: 3.896 min
Delta R.T.: 0.000 min
Response: 2489739
Conc: 49.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.871 min
Delta R.T.: -0.005 min
Response: 4738859
Conc: 51.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.142 min
Delta R.T.: -0.005 min
Response: 2205855
Conc: 45.46 ng/ml




#26 AR-1254-1

R.T.: 7.130 min
Delta R.T.: -0.002 min
Response: 1769834
Conc: 556.81 ng/ml m

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
4/5/2021 3:32:45 PM




#26 AR-1254-1

R.T.: 6.005 min
Delta R.T.: -0.002 min
Response: 1385408
Conc: 528.55 ng/ml m




#27 AR-1254-2

R.T.: 7.360 min
Delta R.T.: 0.000 min
Response: 2923531
Conc: 570.59 ng/ml



#27 AR-1254-2

R.T.: 6.165 min
Delta R.T.: -0.002 min
Response: 1226111
Conc: 535.42 ng/ml m




#28 AR-1254-3

R.T.: 7.742 min
Delta R.T.: 0.000 min
Response: 3289650
Conc: 596.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
4/5/2021 3:32:45 PM




#28 AR-1254-3

R.T.: 6.582 min
Delta R.T.: -0.002 min
Response: 1980726
Conc: 560.43 ng/ml m




#29 AR-1254-4

R.T.: 8.039 min
Delta R.T.: 0.000 min
Response: 2416853
Conc: 613.84 ng/ml



#29 AR-1254-4

R.T.: 6.823 min
Delta R.T.: -0.002 min
Response: 1229140
Conc: 585.67 ng/ml




#30 AR-1254-5

R.T.: 8.465 min
Delta R.T.: -0.002 min
Response: 2512172
Conc: 571.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Ankita
4/5/2021 3:32:45 PM



#30 AR-1254-5

R.T.: 7.246 min
Delta R.T.: -0.002 min
Response: 1778526
Conc: 545.56 ng/ml