

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0040323.M
 Title : GC EXTRACTABLES
 Last Update : Tue Apr 04 05:27:16 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0093752.D 1000 =P0093750.D 750 =P0093751.D
 250 =P0093753.D 50 =P0093754.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.953	3.806	3.894	3.919	3.682	3.851	E6	2.83
2) SA	Decachlorobiphenyl	2.508	2.300	2.394	2.546	2.227	2.395	E6	5.63
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	1.365	1.245	1.298	1.432	1.442	1.356	E5	6.27
27) L6	AR-1254-2	1.980	1.817	1.887	2.044	2.046	1.955	E5	5.15
28) L6	AR-1254-3	1.876	1.744	1.798	1.913	1.907	1.847	E5	4.00
29) L6	AR-1254-4	1.055	0.958	1.003	1.081	0.974	1.014	E5	5.17
30) L6	AR-1254-5	1.356	1.244	1.292	1.424	1.463	1.356	E5	6.67
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0093752.D 1000 =P0093750.D 750 =P0093751.D
 250 =P0093753.D 50 =P0093754.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.361	1.295	1.328	1.361	1.244	1.318	E6	3.77
2) SA	Decachlorobiphenyl	1.084	0.978	1.024	1.133	1.059	1.056	E6	5.59
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0040323.M
 Title : GC EXTRACTABLES
 Last Update : Tue Apr 04 05:27:16 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0093752.D 1000 =P0093750.D 750 =P0093751.D
 250 =P0093753.D 50 =P0093754.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	6.942	6.288	6.581	7.275	7.091	6.835 E4	5.82
27) L6	AR-1254-2	6.167	5.561	5.810	6.533	6.381	6.091 E4	6.59
28) L6	AR-1254-3	9.315	8.478	8.831	9.597	9.323	9.109 E4	4.92
29) L6	AR-1254-4	4.468	4.039	4.226	4.550	4.015	4.260 E4	5.72
30) L6	AR-1254-5	7.605	6.890	7.178	7.769	6.634	7.215 E4	6.59
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range