

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0040323.M
 Title : GC EXTRACTABLES
 Last Update : Tue Apr 04 05:54:34 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0093758.D 1000 =P0093756.D 750 =P0093757.D
 250 =P0093759.D 50 =P0093760.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.930	3.816	3.829	3.976	3.516	3.813	E6	4.71
2) SA	Decachlorobiphenyl	4.325	4.051	4.122	4.479	4.057	4.207	E6	4.48
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	3.451	3.305	3.333	3.537	3.202	3.366	E5	3.87
42) L9	AR-1268-2	3.307	3.052	3.113	3.540	3.079	3.218	E5	6.40
43) L9	AR-1268-3	2.906	2.751	2.782	2.993	2.748	2.836	E5	3.85
44) L9	AR-1268-4	1.013	0.967	0.976	1.034	0.845	0.967	E5	7.59
45) L9	AR-1268-5	8.611	8.332	8.317	8.750	7.983	8.399	E5	3.53

Signal #2 Calibration Files

500 =P0093758.D 1000 =P0093756.D 750 =P0093757.D
 250 =P0093759.D 50 =P0093760.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.352	1.299	1.311	1.382	1.192	1.307	E6	5.53
2) SA	Decachlorobiphenyl	1.814	1.689	1.714	1.935	1.845	1.799	E6	5.56
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0040323.M
 Title : GC EXTRACTABLES
 Last Update : Tue Apr 04 05:54:34 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0093758.D 1000 =P0093756.D 750 =P0093757.D
 250 =P0093759.D 50 =P0093760.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.542	1.449	1.464	1.617	1.566	1.528 E5	4.62
42) L9	AR-1268-2	1.388	1.321	1.331	1.444	1.308	1.359 E5	4.17
43) L9	AR-1268-3	1.327	1.243	1.261	1.393	1.320	1.309 E5	4.57
44) L9	AR-1268-4	4.187	3.847	3.916	4.350	3.492	3.958 E4	8.36
45) L9	AR-1268-5	3.578	3.402	3.429	3.656	3.459	3.505 E5	3.08

(#) = Out of Range