

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040323\
 Data File : P0093754.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2023 00:36
 Operator : YP/AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 05:25:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 05:14:09 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.421	3.625	18411650	6217672	4.781	4.718
2)	SA Decachlor...	10.248	8.648	11135363	5295044	4.650	5.016
Target Compounds							
26)	L6 AR-1254-1	6.460	5.512	7211161	3545358	53.169	51.868
27)	L6 AR-1254-2	6.681	5.660	10230640	3190294	52.338	52.381
28)	L6 AR-1254-3	7.050	6.064	9532510	4661556	51.599	51.176
29)	L6 AR-1254-4	7.337	6.293	4868973	2007504	48.012	47.130
30)	L6 AR-1254-5	7.762	6.714	8578622	3317073	62.962	45.973 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0040323\
Data File : P0093754.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2023 00:36
Operator : YP/AJ
Sample : AR1254IC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_0
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 05:25:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 05:14:09 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.421 min
Delta R.T.: 0.000 min
Response: 18411650
Conc: 4.78 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.625 min
Delta R.T.: 0.000 min
Response: 6217672
Conc: 4.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.248 min
Delta R.T.: 0.000 min
Response: 11135363
Conc: 4.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.648 min
Delta R.T.: 0.000 min
Response: 5295044
Conc: 5.02 ng/ml



#26 AR-1254-1


R.T.: 6.460 min
Delta R.T.: 0.000 min
Response: 7211161
Conc: 53.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 5.512 min
Delta R.T.: 0.000 min
Response: 3545358
Conc: 51.87 ng/ml




#27 AR-1254-2

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 10230640
Conc: 52.34 ng/ml



#27 AR-1254-2


R.T.: 5.660 min
Delta R.T.: 0.000 min
Response: 3190294
Conc: 52.38 ng/ml



#28 AR-1254-3


R.T.: 7.050 min
Delta R.T.: 0.000 min
Response: 9532510
Conc: 51.60 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 6.064 min
Delta R.T.: 0.000 min
Response: 4661556
Conc: 51.18 ng/ml




#29 AR-1254-4

R.T.: 7.337 min
Delta R.T.: 0.000 min
Response: 4868973
Conc: 48.01 ng/ml



#29 AR-1254-4


R.T.: 6.293 min
Delta R.T.: 0.000 min
Response: 2007504
Conc: 47.13 ng/ml



#30 AR-1254-5

R.T.: 7.762 min
Delta R.T.: 0.000 min
Response: 8578622
Conc: 62.96 ng/ml

Instrument :
ECD_O
ClientSampleId :



#30 AR-1254-5

R.T.: 6.714 min
Delta R.T.: 0.000 min
Response: 3317073
Conc: 45.97 ng/ml