

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040323\
 Data File : P0093735.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2023 19:11
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 04:16:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 04:16:16 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.422	3.627	190.5E6	64564128	50.000	50.000
2) SA Decachlor...	10.251	8.651	118.5E6	50285377	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.599	4.710	51624908	18053846	500.000	507.882
4) L1 AR-1016-2	5.621	4.729	78594443	27612922	500.000	500.000
5) L1 AR-1016-3	5.683	4.905	50859563	14697469	500.000	500.000
6) L1 AR-1016-4	5.782	4.948	37783647	13547085	500.000	500.000
7) L1 AR-1016-5	6.080	5.161	42423686	16687717	500.000	500.000
31) L7 AR-1260-1	7.217	6.197	73731347	33054136	500.000	500.000
32) L7 AR-1260-2	7.477	6.387	78447501	36644709	500.000	500.000
33) L7 AR-1260-3	7.840	6.540	66806083	35105902	500.000	500.000
34) L7 AR-1260-4	8.067	7.014	69398793	28170165	500.000	500.000
35) L7 AR-1260-5	8.394	7.258	112.5E6	54032588	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040323\
Data File : PO093735.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2023 19:11
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 04:16:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 04:16:16 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.422 min
Delta R.T.: 0.000 min
Response: 190495565
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.627 min
Delta R.T.: 0.000 min
Response: 64564128
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.251 min
Delta R.T.: 0.000 min
Response: 118464979
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.651 min
Delta R.T.: 0.000 min
Response: 50285377
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.599 min
Delta R.T.: 0.000 min
Response: 51624908
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.710 min
Delta R.T.: 0.000 min
Response: 18053846
Conc: 507.88 ng/ml




#4 AR-1016-2

R.T.: 5.621 min
Delta R.T.: 0.000 min
Response: 78594443
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.729 min
Delta R.T.: 0.000 min
Response: 27612922
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 50859563
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.905 min
Delta R.T.: 0.000 min
Response: 14697469
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.782 min
Delta R.T.: 0.000 min
Response: 37783647
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.948 min
Delta R.T.: 0.000 min
Response: 13547085
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.080 min
Delta R.T.: 0.000 min
Response: 42423686
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.161 min
Delta R.T.: 0.000 min
Response: 16687717
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.217 min
Delta R.T.: 0.000 min
Response: 73731347
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.197 min
Delta R.T.: 0.000 min
Response: 33054136
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.477 min
Delta R.T.: 0.000 min
Response: 78447501
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.387 min
Delta R.T.: 0.000 min
Response: 36644709
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.840 min
Delta R.T.: 0.000 min
Response: 66806083
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.540 min
Delta R.T.: 0.000 min
Response: 35105902
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.067 min
Delta R.T.: 0.000 min
Response: 69398793
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.014 min
Delta R.T.: 0.000 min
Response: 28170165
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.394 min
Delta R.T.: 0.000 min
Response: 112487281
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.258 min
Delta R.T.: 0.000 min
Response: 54032588
Conc: 500.00 ng/ml