

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040324\
 Data File : P0102699.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2024 14:27
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 16:54:03 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 08:56:28 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.264	3.379	212.7E6	141.3E6	48.249	49.098
2)	SA Decachlor...	9.800	8.243	125.7E6	83805191	45.257	46.195
Target Compounds							
16)	L4 AR-1242-1	5.408	4.430	47040073	30143763	522.122	516.460
17)	L4 AR-1242-2	5.430	4.448	68858378	43514460	521.685	533.665
18)	L4 AR-1242-3	5.491	4.619	44125484	23587417	520.409	523.597
19)	L4 AR-1242-4	5.588	4.701	35207772	26294576	524.809	507.318
20)	L4 AR-1242-5	6.313	5.213	30606783	29313621	488.839	523.524



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040324\
Data File : PO102699.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2024 14:27
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 16:54:03 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031824.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 08:56:28 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.264 min
Delta R.T.: -0.003 min
Response: 212725868
Conc: 48.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.379 min
Delta R.T.: -0.003 min
Response: 141270675
Conc: 49.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.800 min
Delta R.T.: -0.009 min
Response: 125654285
Conc: 45.26 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.243 min
Delta R.T.: -0.006 min
Response: 83805191
Conc: 46.19 ng/ml



#16 AR-1242-1


R.T.: 5.408 min
Delta R.T.: -0.004 min
Response: 47040073
Conc: 522.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.430 min
Delta R.T.: -0.003 min
Response: 30143763
Conc: 516.46 ng/ml




#17 AR-1242-2

R.T.: 5.430 min
Delta R.T.: -0.004 min
Response: 68858378
Conc: 521.68 ng/ml



#17 AR-1242-2


R.T.: 4.448 min
Delta R.T.: -0.004 min
Response: 43514460
Conc: 533.66 ng/ml



#18 AR-1242-3


R.T.: 5.491 min
Delta R.T.: -0.004 min
Response: 44125484
Conc: 520.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.619 min
Delta R.T.: -0.004 min
Response: 23587417
Conc: 523.60 ng/ml




#19 AR-1242-4

R.T.: 5.588 min
Delta R.T.: -0.003 min
Response: 35207772
Conc: 524.81 ng/ml



#19 AR-1242-4


R.T.: 4.701 min
Delta R.T.: -0.004 min
Response: 26294576
Conc: 507.32 ng/ml



#20 AR-1242-5

R.T.: 6.313 min
Delta R.T.: -0.005 min
Response: 30606783
Conc: 488.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.213 min
Delta R.T.: -0.004 min
Response: 29313621
Conc: 523.52 ng/ml