

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040324\
 Data File : P0102701.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2024 15:01
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 17:14:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031824.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 19 08:56:28 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.264	3.379	216.6E6	143.1E6	49.122	49.744
2)	SA Decachlor...	9.802	8.244	127.6E6	87861630	45.971	48.431
Target Compounds							
26)	L6 AR-1254-1	6.251	5.212	66341706	63065092	499.388	524.009
27)	L6 AR-1254-2	6.466	5.358	96824160	57181916	508.138	520.007
28)	L6 AR-1254-3	6.829	5.754	93981486	84487340	525.329	535.771
29)	L6 AR-1254-4	7.110	5.980	54866295	44081659	520.703	568.893
30)	L6 AR-1254-5	7.526	6.390	70978684	75398422	527.338	547.122



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040324\
Data File : PO102701.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2024 15:01
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 17:14:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031824.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 19 08:56:28 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.264 min
Delta R.T.: -0.003 min
Response: 216574585
Conc: 49.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.379 min
Delta R.T.: -0.003 min
Response: 143130008
Conc: 49.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.802 min
Delta R.T.: -0.007 min
Response: 127635175
Conc: 45.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.244 min
Delta R.T.: -0.005 min
Response: 87861630
Conc: 48.43 ng/ml



#26 AR-1254-1


R.T.: 6.251 min
Delta R.T.: -0.004 min
Response: 66341706
Conc: 499.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.212 min
Delta R.T.: -0.005 min
Response: 63065092
Conc: 524.01 ng/ml




#27 AR-1254-2

R.T.: 6.466 min
Delta R.T.: -0.004 min
Response: 96824160
Conc: 508.14 ng/ml



#27 AR-1254-2


R.T.: 5.358 min
Delta R.T.: -0.004 min
Response: 57181916
Conc: 520.01 ng/ml



#28 AR-1254-3


R.T.: 6.829 min
Delta R.T.: -0.004 min
Response: 93981486
Conc: 525.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.754 min
Delta R.T.: -0.004 min
Response: 84487340
Conc: 535.77 ng/ml




#29 AR-1254-4

R.T.: 7.110 min
Delta R.T.: -0.004 min
Response: 54866295
Conc: 520.70 ng/ml



#29 AR-1254-4


R.T.: 5.980 min
Delta R.T.: -0.004 min
Response: 44081659
Conc: 568.89 ng/ml



#30 AR-1254-5

R.T.: 7.526 min
Delta R.T.: -0.003 min
Response: 70978684
Conc: 527.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.390 min
Delta R.T.: -0.005 min
Response: 75398422
Conc: 547.12 ng/ml