

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0040418.M
 Title : GC EXTRACTABLES
 Last Update : Wed Apr 04 15:17:57 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0043558.D 1000 =P0043556.D 750 =P0043557.D
 250 =P0043559.D 50 =P0043560.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		1.801	1.666	1.696	1.834	1.893	1.778	E5	5.35
2) SA	Decachlorobiphenyl		1.815	1.615	1.729	1.876	1.975	1.802	E5	7.64
3) L1	AR-1016-1		4.427	3.947	4.158	4.649	4.865	4.409	E3	8.35
4) L1	AR-1016-2		5.871	5.159	5.507	6.119	6.949	5.921	E3	11.49
5) L1	AR-1016-3		4.922	4.393	4.620	5.108	5.092	4.827	E3	6.46
6) L1	AR-1016-4		5.564	5.053	5.303	5.722	6.412	5.611	E3	9.18
7) L1	AR-1016-5		4.554	5.344	4.682	4.636	4.508	4.745	E3	7.20
8) L2	AR-1221-1		3.265					3.265	E3	0.00
9) L2	AR-1221-2		2.376					2.376	E3	0.00
10) L2	AR-1221-3		7.758					7.758	E3	0.00
11) L3	AR-1232-1		3.983					3.983	E3	0.00
12) L3	AR-1232-2		2.915					2.915	E3	0.00
13) L3	AR-1232-3		4.799					4.799	E3	0.00
14) L3	AR-1232-4		2.806					2.806	E3	0.00
15) L3	AR-1232-5		2.215					2.215	E3	0.00
16) L4	AR-1242-1		3.444					3.444	E3	0.00
17) L4	AR-1242-2		7.988					7.988	E3	0.00
18) L4	AR-1242-3		3.856					3.856	E3	0.00
19) L4	AR-1242-4		3.953					3.953	E3	0.00
20) L4	AR-1242-5		4.620					4.620	E3	0.00
21) L5	AR-1248-1		7.023					7.023	E3	0.00
22) L5	AR-1248-2		6.183					6.183	E3	0.00
23) L5	AR-1248-3		8.717					8.717	E3	0.00
24) L5	AR-1248-4		8.962					8.962	E3	0.00
25) L5	AR-1248-5		5.631					5.631	E3	0.00
26) L6	AR-1254-1		1.500					1.500	E4	0.00
27) L6	AR-1254-2		9.495					9.495	E3	0.00
28) L6	AR-1254-3		1.650					1.650	E4	0.00
29) L6	AR-1254-4		1.362					1.362	E4	0.00
30) L6	AR-1254-5		9.754					9.754	E3	0.00
31) L7	AR-1260-1		1.002	0.899	0.942	1.065	1.129	1.007	E4	9.18
32) L7	AR-1260-2		1.293	1.143	1.196	1.362	1.422	1.283	E4	8.98
33) L7	AR-1260-3		1.017	0.915	0.992	1.119	1.011	1.011	E4	7.21
34) L7	AR-1260-4		2.146	1.925	2.120	1.819	2.334	2.069	E4	9.71
35) L7	AR-1260-5		1.237	1.201	1.267	1.226	1.328	1.252	E4	3.88
36) L8	AR-1262-1		9.612					9.612	E3	0.00
37) L8	AR-1262-2		1.225					1.225	E4	0.00
38) L8	AR-1262-3		1.127					1.127	E4	0.00
39) L8	AR-1262-4		1.496					1.496	E4	0.00
40) L8	AR-1262-5		3.043					3.043	E4	0.00
41) L9	AR-1268-1		3.148					3.148	E4	0.00
42) L9	AR-1268-2		2.786					2.786	E4	0.00
43) L9	AR-1268-3		2.414					2.414	E4	0.00
44) L9	AR-1268-4		9.866					9.866	E3	0.00
45) L9	AR-1268-5		6.881					6.881	E4	0.00

Signal #2 Calibration Files

500 =P0043558.D 1000 =P0043556.D 750 =P0043557.D
 250 =P0043559.D 50 =P0043560.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		2.686	2.512	2.489	2.769	2.977	2.687	E5	7.45
2) SA	Decachlorobiphenyl		2.404	1.837	2.321	2.514	2.647	2.344	E5	13.18

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0040418.M
 Title : GC EXTRACTABLES
 Last Update : Wed Apr 04 15:17:57 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0043558.D 1000 =P0043556.D 750 =P0043557.D
 250 =P0043559.D 50 =P0043560.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	7.121	6.162	6.363	6.817	7.629	6.818	E3	8.63
4)	L1	AR-1016-2	6.081	5.132	5.583	6.584	5.756	5.827	E3	9.34
5)	L1	AR-1016-3	7.346	6.233	6.748	7.798	7.151	7.055	E3	8.43
6)	L1	AR-1016-4	7.638	6.730	6.999	8.060	9.172	7.720	E3	12.51
7)	L1	AR-1016-5	7.485	6.920	6.595	7.663	8.571	7.447	E3	10.21
8)	L2	AR-1221-1	4.891					4.891	E3	0.00
9)	L2	AR-1221-2	3.504					3.504	E3	0.00
10)	L2	AR-1221-3	1.120					1.120	E4	0.00
11)	L3	AR-1232-1	5.860					5.860	E3	0.00
12)	L3	AR-1232-2	1.194					1.194	E4	0.00
13)	L3	AR-1232-3	5.199					5.199	E3	0.00
14)	L3	AR-1232-4	4.039					4.039	E3	0.00
15)	L3	AR-1232-5	2.935					2.935	E3	0.00
16)	L4	AR-1242-1	5.873					5.873	E3	0.00
17)	L4	AR-1242-2	2.058					2.058	E4	0.00
18)	L4	AR-1242-3	6.341					6.341	E3	0.00
19)	L4	AR-1242-4	6.106					6.106	E3	0.00
20)	L4	AR-1242-5	6.431					6.431	E3	0.00
21)	L5	AR-1248-1	1.156					1.156	E4	0.00
22)	L5	AR-1248-2	1.060					1.060	E4	0.00
23)	L5	AR-1248-3	1.755					1.755	E4	0.00
24)	L5	AR-1248-4	1.402					1.402	E4	0.00
25)	L5	AR-1248-5	9.747					9.747	E3	0.00
26)	L6	AR-1254-1	2.176					2.176	E4	0.00
27)	L6	AR-1254-2	1.938					1.938	E4	0.00
28)	L6	AR-1254-3	3.166					3.166	E4	0.00
29)	L6	AR-1254-4	2.468					2.468	E4	0.00
30)	L6	AR-1254-5	1.449					1.449	E4	0.00
31)	L7	AR-1260-1	1.682	1.477	1.374	1.876	1.947	1.671	E4	14.78
32)	L7	AR-1260-2	2.145	1.852	1.807	2.288	2.488	2.116	E4	13.66
33)	L7	AR-1260-3	2.147	1.917	2.046	2.325	2.575	2.202	E4	11.63
34)	L7	AR-1260-4	3.391	3.548	3.343	4.043	4.145	3.694	E4	10.14
35)	L7	AR-1260-5	2.411	2.739	2.321	2.677	2.710	2.572	E4	7.44
36)	L8	AR-1262-1	2.176					2.176	E4	0.00
37)	L8	AR-1262-2	2.018					2.018	E4	0.00
38)	L8	AR-1262-3	3.483					3.483	E4	0.00
39)	L8	AR-1262-4	2.712					2.712	E4	0.00
40)	L8	AR-1262-5	5.831					5.831	E4	0.00
41)	L9	AR-1268-1	5.824					5.824	E4	0.00
42)	L9	AR-1268-2	5.219					5.219	E4	0.00
43)	L9	AR-1268-3	4.406					4.406	E4	0.00
44)	L9	AR-1268-4	1.636					1.636	E4	0.00
45)	L9	AR-1268-5	1.106					1.106	E5	0.00

(#) = Out of Range