

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040423\
 Data File : P0093798.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2023 18:38
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 22:06:46 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 06:20:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.422	3.627	187.4E6	65787144	50.128	51.662
2) SA Decachlor...	10.248	8.649	122.2E6	53700481	54.639	53.712
Target Compounds						
26) L6 AR-1254-1	6.461	5.513	61779618	31852595	455.512	465.998
27) L6 AR-1254-2	6.681	5.661	89865261	28969367	459.732	475.646
28) L6 AR-1254-3	7.050	6.066	84719897	43198022	458.583	474.245
29) L6 AR-1254-4	7.338	6.295	52465812	21516439	517.352	505.137
30) L6 AR-1254-5	7.761	6.714	67126937	38282626	495.020	530.574



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040423\
Data File : PO093798.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2023 18:38
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 22:06:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 06:20:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.422 min
Delta R.T.: 0.000 min
Response: 187390554
Conc: 50.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.627 min
Delta R.T.: 0.000 min
Response: 65787144
Conc: 51.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.248 min
Delta R.T.: -0.003 min
Response: 122224145
Conc: 54.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.649 min
Delta R.T.: -0.002 min
Response: 53700481
Conc: 53.71 ng/ml



#26 AR-1254-1


R.T.: 6.461 min
Delta R.T.: 0.000 min
Response: 61779618
Conc: 455.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.513 min
Delta R.T.: 0.000 min
Response: 31852595
Conc: 466.00 ng/ml




#27 AR-1254-2

R.T.: 6.681 min
Delta R.T.: 0.000 min
Response: 89865261
Conc: 459.73 ng/ml



#27 AR-1254-2


R.T.: 5.661 min
Delta R.T.: 0.000 min
Response: 28969367
Conc: 475.65 ng/ml



#28 AR-1254-3


R.T.: 7.050 min
Delta R.T.: 0.000 min
Response: 84719897
Conc: 458.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.066 min
Delta R.T.: 0.000 min
Response: 43198022
Conc: 474.24 ng/ml




#29 AR-1254-4

R.T.: 7.338 min
Delta R.T.: 0.000 min
Response: 52465812
Conc: 517.35 ng/ml



#29 AR-1254-4


R.T.: 6.295 min
Delta R.T.: 0.000 min
Response: 21516439
Conc: 505.14 ng/ml



#30 AR-1254-5

R.T.: 7.761 min
Delta R.T.: 0.000 min
Response: 67126937
Conc: 495.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.714 min
Delta R.T.: 0.000 min
Response: 38282626
Conc: 530.57 ng/ml