

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040523\
 Data File : P0093841.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Apr 2023 19:00
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 22:37:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 06:20:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.420	3.625	193.5E6	66905580	51.750	52.540
2) SA Decachlor...	10.241	8.644	126.8E6	53776340	56.699	53.788
Target Compounds						
21) L5 AR-1248-1	5.595	4.707	34533773	12887422	513.268	540.318
22) L5 AR-1248-2	5.872	4.945	53142570	19117621	485.823	479.943
23) L5 AR-1248-3	6.077	4.986	59019664	20769794	488.896	497.611
24) L5 AR-1248-4	6.484	5.158	57177823	23509136	532.395	499.938
25) L5 AR-1248-5	6.523	5.550	59326124	21335046	511.327	571.465



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040523\
Data File : PO093841.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Apr 2023 19:00
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 22:37:51 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 06:20:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.420 min
Delta R.T.: -0.002 min
Response: 193453427
Conc: 51.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.625 min
Delta R.T.: -0.002 min
Response: 66905580
Conc: 52.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.241 min
Delta R.T.: -0.010 min
Response: 126831214
Conc: 56.70 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.644 min
Delta R.T.: -0.008 min
Response: 53776340
Conc: 53.79 ng/ml



#21 AR-1248-1


R.T.: 5.595 min
Delta R.T.: -0.002 min
Response: 34533773
Conc: 513.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.707 min
Delta R.T.: -0.002 min
Response: 12887422
Conc: 540.32 ng/ml




#22 AR-1248-2

R.T.: 5.872 min
Delta R.T.: -0.001 min
Response: 53142570
Conc: 485.82 ng/ml



#22 AR-1248-2


R.T.: 4.945 min
Delta R.T.: -0.002 min
Response: 19117621
Conc: 479.94 ng/ml



#23 AR-1248-3


R.T.: 6.077 min
Delta R.T.: -0.002 min
Response: 59019664
Conc: 488.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.986 min
Delta R.T.: -0.002 min
Response: 20769794
Conc: 497.61 ng/ml




#24 AR-1248-4

R.T.: 6.484 min
Delta R.T.: -0.002 min
Response: 57177823
Conc: 532.40 ng/ml



#24 AR-1248-4


R.T.: 5.158 min
Delta R.T.: -0.002 min
Response: 23509136
Conc: 499.94 ng/ml



#25 AR-1248-5

R.T.: 6.523 min
Delta R.T.: -0.002 min
Response: 59326124
Conc: 511.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.550 min
Delta R.T.: -0.003 min
Response: 21335046
Conc: 571.47 ng/ml