

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040623\
 Data File : P0093891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Apr 2023 02:00
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 04:56:40 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 06:20:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.420	3.625	177.1E6	63746000	47.387	50.059
2) SA Decachlor...	10.243	8.645	118.0E6	50480615	52.744	50.491
Target Compounds						
16) L4 AR-1242-1	5.596	4.708	42441479	15979076	486.244	519.263
17) L4 AR-1242-2	5.618	4.726	62580893	23429453	470.656	509.402
18) L4 AR-1242-3	5.681	4.902	40877389	12318594	481.547	508.314
19) L4 AR-1242-4	5.780	4.986	30875473	13849570	494.157	493.329
20) L4 AR-1242-5	6.523	5.509	30691134	15008484	455.830	526.412



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040623\
Data File : PO093891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Apr 2023 02:00
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 04:56:40 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 06:20:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.420 min
Delta R.T.: -0.002 min
Response: 177143757
Conc: 47.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.625 min
Delta R.T.: -0.002 min
Response: 63746000
Conc: 50.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.243 min
Delta R.T.: -0.008 min
Response: 117984617
Conc: 52.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.645 min
Delta R.T.: -0.007 min
Response: 50480615
Conc: 50.49 ng/ml



#16 AR-1242-1


R.T.: 5.596 min
Delta R.T.: -0.002 min
Response: 42441479
Conc: 486.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.708 min
Delta R.T.: -0.002 min
Response: 15979076
Conc: 519.26 ng/ml




#17 AR-1242-2

R.T.: 5.618 min
Delta R.T.: -0.001 min
Response: 62580893
Conc: 470.66 ng/ml



#17 AR-1242-2


R.T.: 4.726 min
Delta R.T.: -0.002 min
Response: 23429453
Conc: 509.40 ng/ml



#18 AR-1242-3


R.T.: 5.681 min
Delta R.T.: -0.002 min
Response: 40877389
Conc: 481.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.902 min
Delta R.T.: -0.002 min
Response: 12318594
Conc: 508.31 ng/ml




#19 AR-1242-4

R.T.: 5.780 min
Delta R.T.: -0.002 min
Response: 30875473
Conc: 494.16 ng/ml



#19 AR-1242-4


R.T.: 4.986 min
Delta R.T.: -0.002 min
Response: 13849570
Conc: 493.33 ng/ml



#20 AR-1242-5

R.T.: 6.523 min
Delta R.T.: -0.002 min
Response: 30691134
Conc: 455.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.509 min
Delta R.T.: -0.002 min
Response: 15008484
Conc: 526.41 ng/ml