

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041023\
 Data File : P0094053.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2023 18:28
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 01:46:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 06:20:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.420	3.624	164.7E6	60878592	44.064	47.807
2) SA Decachlor...	10.241	8.643	118.8E6	52836630	53.094	52.848
Target Compounds						
21) L5 AR-1248-1	5.595	4.706	30465038	11905493	452.795	499.149
22) L5 AR-1248-2	5.872	4.943	48427256	19025087	442.716	477.620
23) L5 AR-1248-3	6.076	4.985	53577404	19836895	443.814	475.260
24) L5 AR-1248-4	6.484	5.157	50008928	22666647	465.644	482.022
25) L5 AR-1248-5	6.523	5.549	52954287	18788116	456.408	503.245



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041023\
Data File : PO094053.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2023 18:28
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 01:46:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 06:20:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.420 min
Delta R.T.: -0.002 min
Response: 164720421
Conc: 44.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: -0.003 min
Response: 60878592
Conc: 47.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.241 min
Delta R.T.: -0.009 min
Response: 118766099
Conc: 53.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.643 min
Delta R.T.: -0.009 min
Response: 52836630
Conc: 52.85 ng/ml



#21 AR-1248-1


R.T.: 5.595 min
Delta R.T.: -0.002 min
Response: 30465038
Conc: 452.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.706 min
Delta R.T.: -0.004 min
Response: 11905493
Conc: 499.15 ng/ml




#22 AR-1248-2

R.T.: 5.872 min
Delta R.T.: -0.002 min
Response: 48427256
Conc: 442.72 ng/ml



#22 AR-1248-2


R.T.: 4.943 min
Delta R.T.: -0.004 min
Response: 19025087
Conc: 477.62 ng/ml



#23 AR-1248-3


R.T.: 6.076 min
Delta R.T.: -0.003 min
Response: 53577404
Conc: 443.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.985 min
Delta R.T.: -0.004 min
Response: 19836895
Conc: 475.26 ng/ml




#24 AR-1248-4

R.T.: 6.484 min
Delta R.T.: -0.002 min
Response: 50008928
Conc: 465.64 ng/ml



#24 AR-1248-4


R.T.: 5.157 min
Delta R.T.: -0.003 min
Response: 22666647
Conc: 482.02 ng/ml



#25 AR-1248-5

R.T.: 6.523 min
Delta R.T.: -0.002 min
Response: 52954287
Conc: 456.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.549 min
Delta R.T.: -0.004 min
Response: 18788116
Conc: 503.24 ng/ml