

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041023\
 Data File : P0094071.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2023 00:34
 Operator : YP/AJ
 Sample : 02264-02
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 02-A-02-B-02-C

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 01:57:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 06:20:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.418	3.622	81356535	29692243	21.763	23.317
2) SA Decachlor...	10.239	8.640	60811476	27402003	27.185	27.408

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041023\
Data File : PO094071.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2023 00:34
Operator : YP/AJ
Sample : 02264-02
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 01:57:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 06:20:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.418 min
Delta R.T.: -0.004 min
Response: 81356535
Conc: 21.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: -0.004 min
Response: 29692243
Conc: 23.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.239 min
Delta R.T.: -0.011 min
Response: 60811476
Conc: 27.19 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.640 min
Delta R.T.: -0.011 min
Response: 27402003
Conc: 27.41 ng/ml