

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041123\
 Data File : P0094105.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Apr 2023 20:06
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 02:43:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 06:20:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.418	3.623	178.4E6	63256984	47.714	49.675
2) SA Decachlor...	10.245	8.643	123.8E6	54060415	55.339	54.072
Target Compounds						
16) L4 AR-1242-1	5.595	4.706	41267899	15287551	472.799	496.791
17) L4 AR-1242-2	5.617	4.724	61195143	23191159	460.234	504.221
18) L4 AR-1242-3	5.680	4.900	39264185	11542201	462.543	476.277
19) L4 AR-1242-4	5.779	4.984	29591272	13523089	473.604	481.700
20) L4 AR-1242-5	6.522	5.507	32009657	14841217	475.413	520.546



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041123\
Data File : PO094105.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Apr 2023 20:06
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 12 02:43:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 06:20:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.418 min
Delta R.T.: -0.004 min
Response: 178366738
Conc: 47.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.623 min
Delta R.T.: -0.004 min
Response: 63256984
Conc: 49.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.245 min
Delta R.T.: -0.006 min
Response: 123789938
Conc: 55.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.643 min
Delta R.T.: -0.008 min
Response: 54060415
Conc: 54.07 ng/ml



#16 AR-1242-1


R.T.: 5.595 min
Delta R.T.: -0.004 min
Response: 41267899
Conc: 472.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.706 min
Delta R.T.: -0.004 min
Response: 15287551
Conc: 496.79 ng/ml




#17 AR-1242-2

R.T.: 5.617 min
Delta R.T.: -0.003 min
Response: 61195143
Conc: 460.23 ng/ml



#17 AR-1242-2


R.T.: 4.724 min
Delta R.T.: -0.004 min
Response: 23191159
Conc: 504.22 ng/ml



#18 AR-1242-3


R.T.: 5.680 min
Delta R.T.: -0.003 min
Response: 39264185
Conc: 462.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.900 min
Delta R.T.: -0.004 min
Response: 11542201
Conc: 476.28 ng/ml




#19 AR-1242-4

R.T.: 5.779 min
Delta R.T.: -0.003 min
Response: 29591272
Conc: 473.60 ng/ml



#19 AR-1242-4


R.T.: 4.984 min
Delta R.T.: -0.004 min
Response: 13523089
Conc: 481.70 ng/ml



#20 AR-1242-5

R.T.: 6.522 min
Delta R.T.: -0.003 min
Response: 32009657
Conc: 475.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.507 min
Delta R.T.: -0.004 min
Response: 14841217
Conc: 520.55 ng/ml